

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122322\
 Data File : PO091609.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Dec 2022 13:54
 Operator : YP/AJ
 Sample : N6162-04 10X
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 BC260571LOZ-1-2

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 27 00:51:12 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 22 03:53:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.293	3.470	4450941	1594842	1.638	1.573
2) SA Decachlor...	10.052	8.477	6255127	2043676	3.324	2.343 #

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122322\
Data File : PO091609.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Dec 2022 13:54
Operator : YP/AJ
Sample : N6162-04 10X
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
BC260571LOZ-1-2

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 27 00:51:12 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122122.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 22 03:53:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.293 min
Delta R.T.: 0.000 min
Response: 4450941
Conc: 1.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC260571LOZ-1-2




#1 Tetrachloro-m-xylene

R.T.: 3.470 min
Delta R.T.: 0.002 min
Response: 1594842
Conc: 1.57 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.052 min
Delta R.T.: 0.020 min
Response: 6255127
Conc: 3.32 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.477 min
Delta R.T.: 0.012 min
Response: 2043676
Conc: 2.34 ng/ml