

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122322\
 Data File : PO091616.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Dec 2022 16:53
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 27 00:54:31 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 22 03:53:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.295	3.470	139.4E6	50357777	51.302	49.660
2) SA Decachlor...	10.051	8.476	94595380	42381720	50.271	48.579
Target Compounds						
21) L5 AR-1248-1	5.472	4.555	27416926	10591627	455.867	471.333
22) L5 AR-1248-2	5.749	4.793	38691146	15022540	461.668	465.510
23) L5 AR-1248-3	5.954	4.834	41166653	15400079	455.845	467.031
24) L5 AR-1248-4	6.362	5.006	43351688	17841023	451.434	465.300
25) L5 AR-1248-5	6.399	5.400	41730624	16435196	453.180	482.138



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122322\
Data File : PO091616.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Dec 2022 16:53
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 27 00:54:31 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122122.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 22 03:53:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.295 min
Delta R.T.: 0.001 min
Response: 139380923
Conc: 51.30 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.470 min
Delta R.T.: 0.001 min
Response: 50357777
Conc: 49.66 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.051 min
Delta R.T.: 0.019 min
Response: 94595380
Conc: 50.27 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.476 min
Delta R.T.: 0.011 min
Response: 42381720
Conc: 48.58 ng/ml



#21 AR-1248-1


R.T.: 5.472 min
Delta R.T.: 0.004 min
Response: 27416926
Conc: 455.87 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.555 min
Delta R.T.: 0.003 min
Response: 10591627
Conc: 471.33 ng/ml




#22 AR-1248-2

R.T.: 5.749 min
Delta R.T.: 0.004 min
Response: 38691146
Conc: 461.67 ng/ml



#22 AR-1248-2


R.T.: 4.793 min
Delta R.T.: 0.004 min
Response: 15022540
Conc: 465.51 ng/ml



#23 AR-1248-3


R.T.: 5.954 min
Delta R.T.: 0.006 min
Response: 41166653
Conc: 455.85 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.834 min
Delta R.T.: 0.004 min
Response: 15400079
Conc: 467.03 ng/ml




#24 AR-1248-4

R.T.: 6.362 min
Delta R.T.: 0.006 min
Response: 43351688
Conc: 451.43 ng/ml



#24 AR-1248-4


R.T.: 5.006 min
Delta R.T.: 0.004 min
Response: 17841023
Conc: 465.30 ng/ml



#25 AR-1248-5

R.T.: 6.399 min
Delta R.T.: 0.006 min
Response: 41730624
Conc: 453.18 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.400 min
Delta R.T.: 0.004 min
Response: 16435196
Conc: 482.14 ng/ml