

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122420\
 Data File : PO074200.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Dec 2020 21:21
 Operator : DD\AJ
 Sample : L5212-07
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 S5-07-A

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 25 02:18:21 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122320.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 24 01:16:32 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.934	3.991	2147463	1246274	21.061	21.723
2) SA Decachlor...	10.960	9.252	2705192	1375365	27.778	25.583
Target Compounds						
26) L6 AR-1254-1	7.172	6.105	127.1E6	97888118	32918.596	28301.837
27) L6 AR-1254-2	7.403	6.264	221.5E6	108.8E6	36860.959	35832.987
28) L6 AR-1254-3	7.787	6.681	285.3E6	206.3E6	47487.471	42638.154
29) L6 AR-1254-4	8.083	6.920	277.4E6	165.0E6	52900.601	46381.338
30) L6 AR-1254-5	8.514	7.346	397.9E6	318.0E6	77926.403	69298.561



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122420\
Data File : PO074200.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 Dec 2020 21:21
Operator : DD\AJ
Sample : L5212-07
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
S5-07-A

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 25 02:18:21 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122320.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 24 01:16:32 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.934 min
Delta R.T.: 0.002 min
Response: 2147463
Conc: 21.06 ng/ml

Instrument :
ECD_O
ClientSampleId :
S5-07-A




#1 Tetrachloro-m-xylene

R.T.: 3.991 min
Delta R.T.: 0.000 min
Response: 1246274
Conc: 21.72 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.960 min
Delta R.T.: 0.008 min
Response: 2705192
Conc: 27.78 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.252 min
Delta R.T.: 0.002 min
Response: 1375365
Conc: 25.58 ng/ml



#26 AR-1254-1


R.T.: 7.172 min
Delta R.T.: 0.004 min
Response: 127078800
Conc: 32918.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
S5-07-A




#26 AR-1254-1

R.T.: 6.105 min
Delta R.T.: 0.002 min
Response: 97888118
Conc: 28301.84 ng/ml




#27 AR-1254-2

R.T.: 7.403 min
Delta R.T.: 0.004 min
Response: 221521203
Conc: 36860.96 ng/ml



#27 AR-1254-2


R.T.: 6.264 min
Delta R.T.: 0.002 min
Response: 108754054
Conc: 35832.99 ng/ml



#28 AR-1254-3


R.T.: 7.787 min
Delta R.T.: 0.005 min
Response: 285281861
Conc: 47487.47 ng/ml

Instrument :
ECD_O
ClientSampleId :
S5-07-A




#28 AR-1254-3

R.T.: 6.681 min
Delta R.T.: 0.002 min
Response: 206331875
Conc: 42638.15 ng/ml




#29 AR-1254-4

R.T.: 8.083 min
Delta R.T.: 0.005 min
Response: 277443029
Conc: 52900.60 ng/ml



#29 AR-1254-4


R.T.: 6.920 min
Delta R.T.: 0.003 min
Response: 165002183
Conc: 46381.34 ng/ml



#30 AR-1254-5

R.T.: 8.514 min
Delta R.T.: 0.006 min
Response: 397895610
Conc: 77926.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
S5-07-A



#30 AR-1254-5

R.T.: 7.346 min
Delta R.T.: 0.003 min
Response: 317996736
Conc: 69298.56 ng/ml