

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0122624\
 Data File : P0108773.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Dec 2024 16:10
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/27/2024
 Supervised By :Ankita Jodhani 12/27/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 26 20:27:36 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0120624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Dec 07 05:58:15 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.703	3.700	486.0E6	272.5E6	55.863	53.753
2) SA Decachlor...	8.766	8.716	305.4E6	183.2E6	41.816	47.280
Target Compounds						
3) L1 AR-1016-1	4.798	4.785	154.5E6	87723512	500.915m	546.652m
4) L1 AR-1016-2	4.819	4.806	223.9E6	119.6E6	536.279	538.453
5) L1 AR-1016-3	4.874	4.981	159.0E6	66516293	543.263m	528.605
6) L1 AR-1016-4	4.996	5.022	127.9E6	54012709	553.059	515.359
7) L1 AR-1016-5	5.254	5.236	133.7E6	74283899	531.341	549.392
31) L7 AR-1260-1	6.296	6.270	210.7E6	119.0E6	460.941	508.557
32) L7 AR-1260-2	6.484	6.457	278.1E6	140.7E6	500.474	501.547
33) L7 AR-1260-3	6.853	6.610	227.1E6	132.0E6	489.769	500.953
34) L7 AR-1260-4	7.113	7.083	190.7E6	106.8E6	448.454	502.137
35) L7 AR-1260-5	7.354	7.322	422.1E6	244.8E6	434.304	505.851

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122624\
Data File : PO108773.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Dec 2024 16:10
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/27/2024
Supervised By :Ankita Jodhani 12/27/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 26 20:27:36 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO120624.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Dec 07 05:58:15 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

