

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0122719\
 Data File : P0065076.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 Dec 2019 7:26
 Operator : SM/AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 30 00:52:26 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0122619.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 26 13:32:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.171	3.446	598600	791461	25.486	28.099
2)	SA Decachlor...	9.680	8.356	463312	578769	18.249	17.667
Target Compounds							
20)	L4 AR-1242-5	6.169	5.321	24493	50288	34.169	50.095 #
24)	L5 AR-1248-4	6.169	0.000	24493	0	19.951	N.D. #
25)	L5 AR-1248-5	6.169	5.321	24493	50288	20.789	25.541
26)	L6 AR-1254-1	6.169	5.321	24493	50288	19.231	22.216
28)	L6 AR-1254-3	6.750	0.000	63781	0	28.527	N.D. #
29)	L6 AR-1254-4	0.000	6.143	0	57344	N.D.	26.551 #
30)	L6 AR-1254-5	0.000	6.482	0	21580	N.D.	7.170 #
31)	L7 AR-1260-1	6.954	0.000	2508	0	1.562	N.D. #
32)	L7 AR-1260-2	0.000	6.143	0	57344	N.D.	20.752 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122719\
Data File : P0065076.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 28 Dec 2019 7:26
Operator : SM/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 30 00:52:26 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122619.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 26 13:32:04 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.171 min
Delta R.T.: -0.006 min
Response: 598600
Conc: 25.49 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.446 min
Delta R.T.: -0.006 min
Response: 791461
Conc: 28.10 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.680 min
Delta R.T.: -0.011 min
Response: 463312
Conc: 18.25 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.356 min
Delta R.T.: -0.008 min
Response: 578769
Conc: 17.67 ng/ml



#20 AR-1242-5


R.T.: 6.169 min
Delta R.T.: -0.067 min
Response: 24493
Conc: 34.17 ng/ml

Instrument :
ECD_O
ClientSampleId :




#20 AR-1242-5

R.T.: 5.321 min
Delta R.T.: 0.017 min
Response: 50288
Conc: 50.09 ng/ml




#24 AR-1248-4

R.T.: 6.169 min
Delta R.T.: -0.029 min
Response: 24493
Conc: 19.95 ng/ml



#24 AR-1248-4


R.T.: 0.000 min
Exp R.T.: 4.955 min
Response: 0
Conc: N.D.



#25 AR-1248-5


R.T.: 6.169 min
Delta R.T.: -0.067 min
Response: 24493
Conc: 20.79 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 5.321 min
Delta R.T.: 0.019 min
Response: 50288
Conc: 25.54 ng/ml




#26 AR-1254-1

R.T.: 6.169 min
Delta R.T.: -0.003 min
Response: 24493
Conc: 19.23 ng/ml



#26 AR-1254-1


R.T.: 5.321 min
Delta R.T.: 0.017 min
Response: 50288
Conc: 22.22 ng/ml



#28 AR-1254-3


R.T.: 6.750 min
Delta R.T.: 0.000 min
Response: 63781
Conc: 28.53 ng/ml

Instrument :
ECD_O
ClientSampleId :




#28 AR-1254-3

R.T.: 0.000 min
Exp R.T. : 5.848 min
Response: 0
Conc: N.D.




#29 AR-1254-4

R.T.: 0.000 min
Exp R.T. : 7.034 min
Response: 0
Conc: N.D.



#29 AR-1254-4


R.T.: 6.143 min
Delta R.T.: 0.066 min
Response: 57344
Conc: 26.55 ng/ml



#30 AR-1254-5


R.T.: 0.000 min
Exp R.T.: 7.449 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#30 AR-1254-5

R.T.: 6.482 min
Delta R.T.: -0.007 min
Response: 21580
Conc: 7.17 ng/ml




#31 AR-1260-1

R.T.: 6.954 min
Delta R.T.: 0.042 min
Response: 2508
Conc: 1.56 ng/ml



#31 AR-1260-1


R.T.: 0.000 min
Exp R.T.: 5.977 min
Response: 0
Conc: N.D.



#32 AR-1260-2

R.T.: 0.000 min
Exp R.T.: 7.168 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :



#32 AR-1260-2

R.T.: 6.143 min
Delta R.T.: -0.021 min
Response: 57344
Conc: 20.75 ng/ml