

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0122719\
 Data File : P0065074.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 Dec 2019 6:53
 Operator : SM/AJ
 Sample : K6443-03
 Misc :
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PH-3

Manual Integrations
 APPROVED

mohammad
 12/30/2019 2:00:12 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 30 00:51:38 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0122619.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 26 13:32:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.172	3.446	681442	912525	29.013	32.397
2) SA Decachlor...	9.683	8.356	479126	588117	18.872	17.952
Target Compounds						
31) L7 AR-1260-1	6.905	5.970	36909	60207	22.987m	27.370
32) L7 AR-1260-2	7.158	6.156	37798	44391	19.545	16.065
33) L7 AR-1260-3	7.517	6.308	30593	46563	21.262	19.033
34) L7 AR-1260-4	7.739	6.774	46045	39725	27.088m	19.704 #
35) L7 AR-1260-5	8.046	7.017	57164	89694	17.690m	19.079

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122719\
Data File : P0065074.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 28 Dec 2019 6:53
Operator : SM/AJ
Sample : K6443-03
Misc :
ALS Vial : 57 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
PH-3


Manual Integrations
APPROVED

mohammad
12/30/2019 2:00:12 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 30 00:51:38 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122619.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 26 13:32:04 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.172 min
Delta R.T.: -0.005 min
Response: 681442
Conc: 29.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
PH-3


Manual Integrations
APPROVED

mohammad
12/30/2019 2:00:12 PM




#1 Tetrachloro-m-xylene

R.T.: 3.446 min
Delta R.T.: -0.006 min
Response: 912525
Conc: 32.40 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.683 min
Delta R.T.: -0.008 min
Response: 479126
Conc: 18.87 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.356 min
Delta R.T.: -0.008 min
Response: 588117
Conc: 17.95 ng/ml




#31 AR-1260-1

R.T.: 6.905 min
Delta R.T.: -0.007 min
Response: 36909
Conc: 22.99 ng/ml m

Instrument :
ECD_O
ClientSampleId :
PH-3


Manual Integrations
APPROVED

mohammad
12/30/2019 2:00:12 PM




#31 AR-1260-1

R.T.: 5.970 min
Delta R.T.: -0.007 min
Response: 60207
Conc: 27.37 ng/ml




#32 AR-1260-2

R.T.: 7.158 min
Delta R.T.: -0.010 min
Response: 37798
Conc: 19.54 ng/ml



#32 AR-1260-2

R.T.: 6.156 min
Delta R.T.: -0.008 min
Response: 44391
Conc: 16.06 ng/ml




#33 AR-1260-3

R.T.: 7.517 min
Delta R.T.: -0.006 min
Response: 30593
Conc: 21.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
PH-3


Manual Integrations
APPROVED

mohammad
12/30/2019 2:00:12 PM




#33 AR-1260-3

R.T.: 6.308 min
Delta R.T.: -0.008 min
Response: 46563
Conc: 19.03 ng/ml




#34 AR-1260-4

R.T.: 7.739 min
Delta R.T.: -0.006 min
Response: 46045
Conc: 27.09 ng/ml m



#34 AR-1260-4

R.T.: 6.774 min
Delta R.T.: -0.008 min
Response: 39725
Conc: 19.70 ng/ml




#35 AR-1260-5

R.T.: 8.046 min
Delta R.T.: -0.007 min
Response: 57164
Conc: 17.69 ng/ml m

Instrument :
ECD_O
ClientSampleId :
PH-3

Manual Integrations
APPROVED

mohammad
12/30/2019 2:00:12 PM



#35 AR-1260-5

R.T.: 7.017 min
Delta R.T.: -0.007 min
Response: 89694
Conc: 19.08 ng/ml