

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122722\
 Data File : P0091625.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Dec 2022 10:53
 Operator : YP/AJ
 Sample : N6185-01
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PPE-1

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/28/2022
 Supervised By :Ankita Jodhani 12/28/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 27 21:13:49 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 22 03:53:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.291	3.467	62508955	27288828	23.008m	26.911m
2) SA Decachlor...	10.033	8.466	51674734	24082547	27.461	27.604

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122722\
Data File : PO091625.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Dec 2022 10:53
Operator : YP/AJ
Sample : N6185-01
Misc :
ALS Vial : 10 Sample Multiplier: 1


Instrument :
ECD_O
ClientSampleId :
PPE-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/28/2022
Supervised By :Ankita Jodhani 12/28/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 27 21:13:49 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122122.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 22 03:53:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.291 min
Delta R.T.: -0.003 min
Response: 62508955
Conc: 23.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
PPE-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/28/2022
Supervised By :Ankita Jodhani 12/28/2022




#1 Tetrachloro-m-xylene

R.T.: 3.467 min
Delta R.T.: -0.001 min
Response: 27288828
Conc: 26.91 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.033 min
Delta R.T.: 0.001 min
Response: 51674734
Conc: 27.46 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.466 min
Delta R.T.: 0.001 min
Response: 24082547
Conc: 27.60 ng/ml