

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0122723\
 Data File : P0100772.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Dec 2023 07:48
 Operator : YP/AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 28 00:13:10 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0122023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 21 04:28:19 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.362	3.661	96723	17130	0.031	0.008 #
2) SA Decachlor...	10.094	8.707	89069	392304	0.043	0.260 #
Target Compounds						
3) L1 AR-1016-1	5.529	4.753	113690	158908	1.325	2.568 #
4) L1 AR-1016-2	5.540	4.753	38308	158908	0.300	1.802 #
5) L1 AR-1016-3	5.597	4.897f	20063	28937	0.254	0.619 #
6) L1 AR-1016-4	5.689	4.983	52912	26996	0.868	0.658
7) L1 AR-1016-5	5.993	5.208	82509	53218	1.348	1.024
8) L2 AR-1221-1	4.550	3.875	194007	181986	5.248	7.247 #
9) L2 AR-1221-2	4.635	3.939	71387	16214	2.571	0.844 #
10) L2 AR-1221-3	4.727	4.036	113399	83002	1.359	1.410
11) L3 AR-1232-1	4.727	4.036	113399	83002	1.701	1.755
12) L3 AR-1232-2	5.255	4.753	38283	158908	1.092	3.810 #
13) L3 AR-1232-3	5.540	4.983f	38308	26996	0.642	1.258 #
14) L3 AR-1232-4	5.689	5.029	52912	28321	1.930	1.333 #
15) L3 AR-1232-5	5.852f	5.208	652344	53218	27.132	2.435 #
16) L4 AR-1242-1	5.529	4.753	113690	158908	1.612	3.043 #
17) L4 AR-1242-2	5.540	4.753	38308	158908	0.365	2.193 #
18) L4 AR-1242-3	5.608	4.983f	19948	26996	0.307	0.705 #
19) L4 AR-1242-4	5.689	5.029	52912	28321	1.057	0.679 #
20) L4 AR-1242-5	6.435	5.568	369293	173598	7.135	3.783 #
21) L5 AR-1248-1	5.529	4.753	113690	158908	2.072	3.974 #
22) L5 AR-1248-2	5.852f	4.983	652344	26996	7.553	0.449 #
23) L5 AR-1248-3	5.993	5.029	82509	28321	0.940	0.453 #
24) L5 AR-1248-4	6.415	5.208	1489305	53218	16.565	0.739 #
25) L5 AR-1248-5	6.435	5.591	369293	88071	4.114	1.428 #
26) L6 AR-1254-1	6.364	5.568	1078597	173598	10.261	1.691 #
27) L6 AR-1254-2	6.594	5.699	131354	448219	0.890	4.879 #
28) L6 AR-1254-3	6.946	6.110	41121	143253	0.294	1.059 #
29) L6 AR-1254-4	0.000	6.323	0	237104	N.D.	3.287 #
30) L6 AR-1254-5	7.626f	6.774	22927492	1242646	193.765	10.785 #
31) L7 AR-1260-1	7.147	6.238	7134383	422466	52.077	4.435 #
32) L7 AR-1260-2	7.385	6.430	66251	82288	0.527	0.766 #
33) L7 AR-1260-3	7.757	6.609	152631	181459	1.576	1.770
34) L7 AR-1260-4	7.968	7.035	118856	113180	1.126	1.388
35) L7 AR-1260-5	8.295	7.304	1662171	732548	7.874	4.059 #
36) L8 AR-1262-1	7.626f	6.774	22927492	1242646	229.505	19.921 #
37) L8 AR-1262-2	8.295	7.035	1662171	113180	7.004	1.002 #
38) L8 AR-1262-3	8.600	7.590	632695	-46285	3.896	N.D. #
39) L8 AR-1262-4	8.682	7.661	279340	713482	2.167	4.830 #
40) L8 AR-1262-5	9.328	8.127	100769	397830	1.263	6.221 #
41) L9 AR-1268-1	8.600	7.590	632695	-46285	2.239	N.D. #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0122723\
 Data File : P0100772.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Dec 2023 07:48
 Operator : YP/AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 28 00:13:10 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0122023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 21 04:28:19 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.682	7.661	279340	713482	1.105	3.528 #
43) L9	AR-1268-3	8.896	7.859	31210	1778379	0.137	9.426 #
44) L9	AR-1268-4	9.328	8.127	100769	397830	1.287	5.925 #
45) L9	AR-1268-5	9.755	8.413	85539	913523	0.119	1.761 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122723\
 Data File : PO100772.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Dec 2023 07:48
 Operator : YP/AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 28 00:13:10 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 21 04:28:19 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.362 min
Delta R.T.: 0.013 min
Response: 96723
Conc: 0.03 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.661 min
Delta R.T.: 0.015 min
Response: 17130
Conc: 0.01 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.094 min
Delta R.T.: 0.002 min
Response: 89069
Conc: 0.04 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.707 min
Delta R.T.: 0.001 min
Response: 392304
Conc: 0.26 ng/ml



#3 AR-1016-1


R.T.: 5.529 min
Delta R.T.: 0.011 min
Response: 113690
Conc: 1.32 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 4.753 min
Delta R.T.: 0.011 min
Response: 158908
Conc: 2.57 ng/ml




#4 AR-1016-2

R.T.: 5.540 min
Delta R.T.: 0.000 min
Response: 38308
Conc: 0.30 ng/ml



#4 AR-1016-2


R.T.: 4.753 min
Delta R.T.: -0.009 min
Response: 158908
Conc: 1.80 ng/ml



#5 AR-1016-3


R.T.: 5.597 min
Delta R.T.: -0.004 min
Response: 20063
Conc: 0.25 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 4.897 min
Delta R.T.: -0.043 min
Response: 28937
Conc: 0.62 ng/ml




#6 AR-1016-4

R.T.: 5.689 min
Delta R.T.: -0.011 min
Response: 52912
Conc: 0.87 ng/ml



#6 AR-1016-4


R.T.: 4.983 min
Delta R.T.: 0.002 min
Response: 26996
Conc: 0.66 ng/ml



#7 AR-1016-5


R.T.: 5.993 min
Delta R.T.: -0.003 min
Response: 82509
Conc: 1.35 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 5.208 min
Delta R.T.: 0.011 min
Response: 53218
Conc: 1.02 ng/ml




#8 AR-1221-1

R.T.: 4.550 min
Delta R.T.: -0.008 min
Response: 194007
Conc: 5.25 ng/ml



#8 AR-1221-1


R.T.: 3.875 min
Delta R.T.: 0.013 min
Response: 181986
Conc: 7.25 ng/ml



#9 AR-1221-2


R.T.: 4.635 min
Delta R.T.: -0.010 min
Response: 71387
Conc: 2.57 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 3.939 min
Delta R.T.: -0.009 min
Response: 16214
Conc: 0.84 ng/ml




#10 AR-1221-3

R.T.: 4.727 min
Delta R.T.: 0.006 min
Response: 113399
Conc: 1.36 ng/ml



#10 AR-1221-3


R.T.: 4.036 min
Delta R.T.: 0.012 min
Response: 83002
Conc: 1.41 ng/ml



#11 AR-1232-1


R.T.: 4.727 min
Delta R.T.: 0.007 min
Response: 113399
Conc: 1.70 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 4.036 min
Delta R.T.: 0.011 min
Response: 83002
Conc: 1.76 ng/ml




#12 AR-1232-2

R.T.: 5.255 min
Delta R.T.: 0.005 min
Response: 38283
Conc: 1.09 ng/ml



#12 AR-1232-2


R.T.: 4.753 min
Delta R.T.: -0.009 min
Response: 158908
Conc: 3.81 ng/ml



#13 AR-1232-3


R.T.: 5.540 min
Delta R.T.: 0.000 min
Response: 38308
Conc: 0.64 ng/ml

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 4.983 min
Delta R.T.: 0.043 min
Response: 26996
Conc: 1.26 ng/ml




#14 AR-1232-4

R.T.: 5.689 min
Delta R.T.: -0.012 min
Response: 52912
Conc: 1.93 ng/ml



#14 AR-1232-4


R.T.: 5.029 min
Delta R.T.: 0.005 min
Response: 28321
Conc: 1.33 ng/ml



#15 AR-1232-5


R.T.: 5.852 min
Delta R.T.: 0.059 min
Response: 652344
Conc: 27.13 ng/ml

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 5.208 min
Delta R.T.: 0.010 min
Response: 53218
Conc: 2.44 ng/ml




#16 AR-1242-1

R.T.: 5.529 min
Delta R.T.: 0.010 min
Response: 113690
Conc: 1.61 ng/ml



#16 AR-1242-1


R.T.: 4.753 min
Delta R.T.: 0.010 min
Response: 158908
Conc: 3.04 ng/ml



#17 AR-1242-2


R.T.: 5.540 min
Delta R.T.: 0.000 min
Response: 38308
Conc: 0.36 ng/ml

Instrument :
ECD_O
ClientSampleId :




#17 AR-1242-2

R.T.: 4.753 min
Delta R.T.: -0.009 min
Response: 158908
Conc: 2.19 ng/ml




#18 AR-1242-3

R.T.: 5.608 min
Delta R.T.: 0.005 min
Response: 19948
Conc: 0.31 ng/ml



#18 AR-1242-3


R.T.: 4.983 min
Delta R.T.: 0.042 min
Response: 26996
Conc: 0.70 ng/ml



#19 AR-1242-4


R.T.: 5.689 min
Delta R.T.: -0.012 min
Response: 52912
Conc: 1.06 ng/ml

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 5.029 min
Delta R.T.: 0.004 min
Response: 28321
Conc: 0.68 ng/ml




#20 AR-1242-5

R.T.: 6.435 min
Delta R.T.: -0.006 min
Response: 369293
Conc: 7.14 ng/ml



#20 AR-1242-5


R.T.: 5.568 min
Delta R.T.: 0.015 min
Response: 173598
Conc: 3.78 ng/ml



#21 AR-1248-1


R.T.: 5.529 min
Delta R.T.: 0.011 min
Response: 113690
Conc: 2.07 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 4.753 min
Delta R.T.: 0.011 min
Response: 158908
Conc: 3.97 ng/ml




#22 AR-1248-2

R.T.: 5.852 min
Delta R.T.: 0.060 min
Response: 652344
Conc: 7.55 ng/ml



#22 AR-1248-2


R.T.: 4.983 min
Delta R.T.: 0.002 min
Response: 26996
Conc: 0.45 ng/ml



#23 AR-1248-3


R.T.: 5.993 min
Delta R.T.: -0.004 min
Response: 82509
Conc: 0.94 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 5.029 min
Delta R.T.: 0.005 min
Response: 28321
Conc: 0.45 ng/ml




#24 AR-1248-4

R.T.: 6.415 min
Delta R.T.: 0.014 min
Response: 1489305
Conc: 16.56 ng/ml



#24 AR-1248-4


R.T.: 5.208 min
Delta R.T.: 0.011 min
Response: 53218
Conc: 0.74 ng/ml



#25 AR-1248-5


R.T.: 6.435 min
Delta R.T.: -0.005 min
Response: 369293
Conc: 4.11 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 5.591 min
Delta R.T.: -0.003 min
Response: 88071
Conc: 1.43 ng/ml




#26 AR-1254-1

R.T.: 6.364 min
Delta R.T.: -0.010 min
Response: 1078597
Conc: 10.26 ng/ml



#26 AR-1254-1


R.T.: 5.568 min
Delta R.T.: 0.016 min
Response: 173598
Conc: 1.69 ng/ml



#27 AR-1254-2


R.T.: 6.594 min
Delta R.T.: 0.000 min
Response: 131354
Conc: 0.89 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 5.699 min
Delta R.T.: -0.002 min
Response: 448219
Conc: 4.88 ng/ml




#28 AR-1254-3

R.T.: 6.946 min
Delta R.T.: -0.015 min
Response: 41121
Conc: 0.29 ng/ml



#28 AR-1254-3


R.T.: 6.110 min
Delta R.T.: 0.001 min
Response: 143253
Conc: 1.06 ng/ml



#29 AR-1254-4


R.T.: 0.000 min
Exp R.T.: 7.247 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 6.323 min
Delta R.T.: -0.016 min
Response: 237104
Conc: 3.29 ng/ml




#30 AR-1254-5

R.T.: 7.626 min
Delta R.T.: -0.042 min
Response: 22927492
Conc: 193.76 ng/ml



#30 AR-1254-5


R.T.: 6.774 min
Delta R.T.: 0.014 min
Response: 1242646
Conc: 10.78 ng/ml



#31 AR-1260-1


R.T.: 7.147 min
Delta R.T.: 0.021 min
Response: 7134383
Conc: 52.08 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 6.238 min
Delta R.T.: -0.001 min
Response: 422466
Conc: 4.44 ng/ml




#32 AR-1260-2

R.T.: 7.385 min
Delta R.T.: 0.001 min
Response: 66251
Conc: 0.53 ng/ml



#32 AR-1260-2


R.T.: 6.430 min
Delta R.T.: 0.001 min
Response: 82288
Conc: 0.77 ng/ml



#33 AR-1260-3


R.T.: 7.757 min
Delta R.T.: 0.012 min
Response: 152631
Conc: 1.58 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 6.609 min
Delta R.T.: 0.025 min
Response: 181459
Conc: 1.77 ng/ml




#34 AR-1260-4

R.T.: 7.968 min
Delta R.T.: -0.002 min
Response: 118856
Conc: 1.13 ng/ml



#34 AR-1260-4


R.T.: 7.035 min
Delta R.T.: -0.023 min
Response: 113180
Conc: 1.39 ng/ml



#35 AR-1260-5


R.T.: 8.295 min
Delta R.T.: 0.006 min
Response: 1662171
Conc: 7.87 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 7.304 min
Delta R.T.: 0.003 min
Response: 732548
Conc: 4.06 ng/ml




#36 AR-1262-1

R.T.: 7.626 min
Delta R.T.: -0.038 min
Response: 22927492
Conc: 229.50 ng/ml



#36 AR-1262-1


R.T.: 6.774 min
Delta R.T.: 0.015 min
Response: 1242646
Conc: 19.92 ng/ml



#37 AR-1262-2


R.T.: 8.295 min
Delta R.T.: 0.006 min
Response: 1662171
Conc: 7.00 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 7.035 min
Delta R.T.: -0.023 min
Response: 113180
Conc: 1.00 ng/ml




#38 AR-1262-3

R.T.: 8.600 min
Delta R.T.: 0.002 min
Response: 632695
Conc: 3.90 ng/ml



#38 AR-1262-3


R.T.: 7.590 min
Delta R.T.: 0.003 min
Response: -46285
Conc: N.D.



#39 AR-1262-4


R.T.: 8.682 min
Delta R.T.: -0.005 min
Response: 279340
Conc: 2.17 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 7.661 min
Delta R.T.: 0.011 min
Response: 713482
Conc: 4.83 ng/ml




#40 AR-1262-5

R.T.: 9.328 min
Delta R.T.: -0.011 min
Response: 100769
Conc: 1.26 ng/ml



#40 AR-1262-5


R.T.: 8.127 min
Delta R.T.: -0.022 min
Response: 397830
Conc: 6.22 ng/ml



#41 AR-1268-1


R.T.: 8.600 min
Delta R.T.: 0.005 min
Response: 632695
Conc: 2.24 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 7.590 min
Delta R.T.: 0.004 min
Response: -46285
Conc: N.D.




#42 AR-1268-2

R.T.: 8.682 min
Delta R.T.: -0.007 min
Response: 279340
Conc: 1.11 ng/ml



#42 AR-1268-2


R.T.: 7.661 min
Delta R.T.: 0.010 min
Response: 713482
Conc: 3.53 ng/ml



#43 AR-1268-3


R.T.: 8.896 min
Delta R.T.: -0.020 min
Response: 31210
Conc: 0.14 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 7.859 min
Delta R.T.: 0.000 min
Response: 1778379
Conc: 9.43 ng/ml




#44 AR-1268-4

R.T.: 9.328 min
Delta R.T.: -0.012 min
Response: 100769
Conc: 1.29 ng/ml



#44 AR-1268-4


R.T.: 8.127 min
Delta R.T.: -0.021 min
Response: 397830
Conc: 5.93 ng/ml



#45 AR-1268-5

R.T.: 9.755 min
Delta R.T.: 0.001 min
Response: 85539
Conc: 0.12 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.413 min
Delta R.T.: -0.032 min
Response: 913523
Conc: 1.76 ng/ml