

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\PO122723\
 Data File : PO100807.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Dec 2023 20:56
 Operator : YP/AJ
 Sample : 06026-01
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 50223

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/28/2023
 Supervised By :Ankita Jodhani 12/28/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 28 04:08:18 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO122023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 21 04:28:19 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.348	3.644	74037346	44658186	23.728	21.855
2) SA Decachlor...	10.091	8.706	49378548	39047479	23.636	25.870
Target Compounds						
16) L4 AR-1242-1	5.517	4.739	31450672	21873069	445.915m	418.798m
17) L4 AR-1242-2	5.538	4.759	39674235	27427630	377.985m	378.448m
18) L4 AR-1242-3	5.601	4.937	26450774	14740257	406.906m	384.698
19) L4 AR-1242-4	5.699	5.021	20342185	11779272	406.216m	282.601 #
20) L4 AR-1242-5	6.440	5.549	1869693	2876140	36.125m	62.668 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122723\
Data File : PO100807.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Dec 2023 20:56
Operator : YP/AJ
Sample : 06026-01
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
50223

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/28/2023
Supervised By :Ankita Jodhani 12/28/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 28 04:08:18 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122023.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 21 04:28:19 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

