

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0122724\
 Data File : P0108814.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Dec 2024 21:18
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/30/2024
 Supervised By :Ankita Jodhani 12/30/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 28 00:30:11 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0120624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Dec 07 05:58:15 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.704	3.700	498.6E6	279.4E6	57.306	55.117
2) SA Decachlor...	8.766	8.716	340.6E6	199.3E6	46.631	51.430
Target Compounds						
3) L1 AR-1016-1	4.799	4.785	164.3E6	84474460	532.576m	526.405m
4) L1 AR-1016-2	4.819	4.806	228.8E6	121.4E6	548.055	546.568
5) L1 AR-1016-3	4.875	4.981	163.0E6	68355504	556.913	543.221
6) L1 AR-1016-4	4.997	5.023	126.6E6	55817442	547.782	532.579
7) L1 AR-1016-5	5.254	5.236	137.7E6	72507226	547.239	536.252
31) L7 AR-1260-1	6.296	6.270	247.9E6	128.9E6	542.250	550.605
32) L7 AR-1260-2	6.484	6.457	299.3E6	154.2E6	538.566	549.759
33) L7 AR-1260-3	6.853	6.610	236.5E6	144.6E6	509.971	548.472
34) L7 AR-1260-4	7.114	7.082	208.1E6	121.4E6	489.186	570.569
35) L7 AR-1260-5	7.354	7.321	465.1E6	273.3E6	478.568	564.723

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122724\
Data File : PO108814.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Dec 2024 21:18
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/30/2024
Supervised By :Ankita Jodhani 12/30/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 28 00:30:11 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO120624.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Dec 07 05:58:15 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

