

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122822\
 Data File : PO091688.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Dec 2022 15:22
 Operator : YP/AJ
 Sample : N6213-07
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 SB-7

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 29 02:05:12 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 22 03:53:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.273	3.470	287.0E6	20179479	105.653	19.900 #
2) SA Decachlor...	10.029	8.463	37111839	16671844	19.722	19.110
Target Compounds						
16) L4 AR-1242-1	5.467	4.553	6043.3E6	2471.3E6	83219.120	89500.562
17) L4 AR-1242-2	5.489	4.571	8374.3E6	3491.0E6	80827.980	90061.843
18) L4 AR-1242-3	5.552	4.747	4910.1E6	1706.5E6	75798.693	81939.930
19) L4 AR-1242-4	5.651	4.832	3810.4E6	1073.1E6	76473.918	48822.733 #
20) L4 AR-1242-5	6.392	5.356	1279.0E6	803.6E6	24453.077	30052.363

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122822\
Data File : PO091688.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Dec 2022 15:22
Operator : YP/AJ
Sample : N6213-07
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
SB-7

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 29 02:05:12 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122122.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 22 03:53:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

