

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\
 Method File : P0122921.M
 Title : GC EXTRACTABLES
 Last Update : Thu Dec 30 06:41:37 2021
 Response Via : Initial Calibration

Calibration Files

500 =P0083950.D 1000 =P0083948.D 750 =P0083949.D
 250 =P0083951.D 50 =P0083952.D

Compound			500	1000	750	250	50	Avg	%RSD	
1)	SA	Tetrachloro-m-xylene	1.347	1.290	1.348	1.322	1.266	1.315	E5	2.73
2)	SA	Decachlorobiphenyl	7.852	7.590	7.879	7.654	7.565	7.708	E4	1.91
3)	L1	AR-1016-1						0.000		-1.00
4)	L1	AR-1016-2						0.000		-1.00
5)	L1	AR-1016-3						0.000		-1.00
6)	L1	AR-1016-4						0.000		-1.00
7)	L1	AR-1016-5						0.000		-1.00
8)	L2	AR-1221-1						0.000		-1.00
9)	L2	AR-1221-2						0.000		-1.00
10)	L2	AR-1221-3						0.000		-1.00
11)	L3	AR-1232-1						0.000		-1.00
12)	L3	AR-1232-2						0.000		-1.00
13)	L3	AR-1232-3						0.000		-1.00
14)	L3	AR-1232-4						0.000		-1.00
15)	L3	AR-1232-5						0.000		-1.00
16)	L4	AR-1242-1	2.653	2.528	2.688	2.584	2.465	2.583	E3	3.51
17)	L4	AR-1242-2	4.485	4.230	4.437	4.539	4.368	4.412	E3	2.72
18)	L4	AR-1242-3	2.649	2.514	2.659	2.621	2.428	2.574	E3	3.88
19)	L4	AR-1242-4	1.993	1.928	2.018	1.943	1.549	1.886	E3	10.18
20)	L4	AR-1242-5	2.047	2.017	2.094	1.984	1.693	1.967	E3	8.06
21)	L5	AR-1248-1						0.000		-1.00
22)	L5	AR-1248-2						0.000		-1.00
23)	L5	AR-1248-3						0.000		-1.00
24)	L5	AR-1248-4						0.000		-1.00
25)	L5	AR-1248-5						0.000		-1.00
26)	L6	AR-1254-1						0.000		-1.00
27)	L6	AR-1254-2						0.000		-1.00
28)	L6	AR-1254-3						0.000		-1.00
29)	L6	AR-1254-4						0.000		-1.00
30)	L6	AR-1254-5						0.000		-1.00
31)	L7	AR-1260-1						0.000		-1.00
32)	L7	AR-1260-2						0.000		-1.00
33)	L7	AR-1260-3						0.000		-1.00
34)	L7	AR-1260-4						0.000		-1.00
35)	L7	AR-1260-5						0.000		-1.00
36)	L8	AR-1262-1						0.000		-1.00
37)	L8	AR-1262-2						0.000		-1.00
38)	L8	AR-1262-3						0.000		-1.00
39)	L8	AR-1262-4						0.000		-1.00
40)	L8	AR-1262-5						0.000		-1.00
41)	L9	AR-1268-1						0.000		-1.00
42)	L9	AR-1268-2						0.000		-1.00
43)	L9	AR-1268-3						0.000		-1.00
44)	L9	AR-1268-4						0.000		-1.00
45)	L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

500 =P0083950.D 1000 =P0083948.D 750 =P0083949.D
 250 =P0083951.D 50 =P0083952.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	2.818	2.686	2.839	2.917	2.740	2.800	E4	3.20
2) SA	Decachlorobiphenyl	1.438	1.427	1.480	1.402	1.297	1.409	E4	4.86
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\
 Method File : P0122921.M
 Title : GC EXTRACTABLES
 Last Update : Thu Dec 30 06:41:37 2021
 Response Via : Initial Calibration

Calibration Files

500 =P0083950.D 1000 =P0083948.D 750 =P0083949.D
 250 =P0083951.D 50 =P0083952.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1	4.204	3.963	4.184	4.099	4.178	4.125 E2	2.41
17) L4	AR-1242-2	8.583	8.009	8.515	8.703	6.679	8.098 E2	10.32
18) L4	AR-1242-3	3.477	3.354	3.532	3.343	3.522	3.446 E2	2.65
19) L4	AR-1242-4	3.292	3.137	3.291	3.178	3.041	3.188 E2	3.36
20) L4	AR-1242-5	3.399	3.284	3.574	3.301	2.738	3.259 E2	9.61
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range