

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0122921.M
 Title : GC EXTRACTABLES
 Last Update : Thu Dec 30 06:48:05 2021
 Response Via : Initial Calibration

Calibration Files

500 =P0083955.D 1000 =P0083953.D 750 =P0083954.D
 250 =P0083956.D 50 =P0083957.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.365	1.317	1.336	1.312	1.284	1.323	E5	2.29
2) SA	Decachlorobiphenyl	8.087	7.737	7.872	7.493	8.048	7.847	E4	3.10
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1	1.985	1.920	1.943	1.935	1.806	1.918	E3	3.49
22) L5	AR-1248-2	3.061	2.875	2.937	2.965	2.786	2.925	E3	3.51
23) L5	AR-1248-3	3.584	3.396	3.487	3.459	3.684	3.522	E3	3.22
24) L5	AR-1248-4	3.506	3.360	3.401	3.310	2.791	3.274	E3	8.52
25) L5	AR-1248-5	3.724	3.552	3.611	3.382	2.627	3.379	E3	12.98
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

500 =P0083955.D 1000 =P0083953.D 750 =P0083954.D
 250 =P0083956.D 50 =P0083957.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.993	2.831	2.897	2.908	2.794	2.884	E4	2.67
2) SA	Decachlorobiphenyl	1.505	1.476	1.491	1.390	1.376	1.448	E4	4.14
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0122921.M
 Title : GC EXTRACTABLES
 Last Update : Thu Dec 30 06:48:05 2021
 Response Via : Initial Calibration

Calibration Files

500 =P0083955.D 1000 =P0083953.D 750 =P0083954.D
 250 =P0083956.D 50 =P0083957.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1	3.299	3.142	3.213	3.114	3.030	3.160 E2	3.22
22) L5	AR-1248-2	5.437	4.994	5.203	5.385	4.808	5.165 E2	5.13
23) L5	AR-1248-3	5.005	4.613	4.767	4.883	3.687	4.591 E2	11.46
24) L5	AR-1248-4	5.784	5.405	5.561	5.680	5.021	5.490 E2	5.42
25) L5	AR-1248-5	4.553	4.527	4.553	4.009	3.919	4.312 E2	7.41
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range