

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0122921.M
 Title : GC EXTRACTABLES
 Last Update : Thu Dec 30 08:27:33 2021
 Response Via : Initial Calibration

Calibration Files

500 =P0083960.D 1000 =P0083958.D 750 =P0083959.D
 250 =P0083961.D 50 =P0083962.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.352	1.288	1.318	1.340	1.276	1.315 E5	2.50
2) SA	Decachlorobiphenyl	7.832	7.589	7.783	7.802	7.601	7.721 E4	1.51
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	4.019	3.751	3.869	4.015	4.051	3.941 E3	3.23
27) L6	AR-1254-2	6.039	5.676	5.852	5.982	6.012	5.912 E3	2.54
28) L6	AR-1254-3	5.891	5.640	5.761	5.693	5.055	5.608 E3	5.77
29) L6	AR-1254-4	3.385	3.274	3.364	3.291	3.492	3.361 E3	2.58
30) L6	AR-1254-5	4.172	4.003	4.097	4.034	4.209	4.103 E3	2.14
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

Signal #2 Calibration Files

500 =P0083960.D 1000 =P0083958.D 750 =P0083959.D
 250 =P0083961.D 50 =P0083962.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	2.931	2.789	2.855	2.938	2.731	2.849 E4	3.16
2) SA	Decachlorobiphenyl	1.466	1.450	1.466	1.391	1.217	1.398 E4	7.54
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0122921.M
 Title : GC EXTRACTABLES
 Last Update : Thu Dec 30 08:27:33 2021
 Response Via : Initial Calibration

Calibration Files

500 =P0083960.D 1000 =P0083958.D 750 =P0083959.D
 250 =P0083961.D 50 =P0083962.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	8.654	7.655	8.122	8.950	8.266	8.329 E2	5.98
27) L6	AR-1254-2	7.716	6.853	7.194	7.796	7.578	7.428 E2	5.33
28) L6	AR-1254-3	1.223	1.119	1.162	1.220	1.261	1.197 E3	4.71
29) L6	AR-1254-4	4.629	4.221	4.401	4.436	4.660	4.469 E2	4.02
30) L6	AR-1254-5	1.026	0.976	1.001	0.989	0.829	0.964 E3	8.08
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range