Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122921\
Data File : PO083947.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Dec 2021 3:49
Operator : AJ\MA
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1232ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 30 04:21:30 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122921.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 30 04:06:00 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1) SA	Tetrachlo...	4.652	3.632	6484130	1350799	50.000	50.000
2) SA	Decachlor...	10.498	8.899	3803761	712959	50.000	50.000
	Target Compounds						
11) L3	AR-1232-1	5.091	4.074	1242092	285430	500.000	500.000
12) L3	AR-1232-2	5.674	4.898	636463	239334	500.000	500.000
13) L3	AR-1232-3	5.995	5.087	1212436	93088	500.000	500.000
14) L3	AR-1232-4	6.167	5.185	526717	76252	500.000	500.000
15) L3	AR-1232-5	6.263	5.367	367813	80084	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122921\
Data File : PO083947.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Dec 2021 3:49
Operator : AJ\MA
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1232ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 30 04:21:30 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122921.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 30 04:06:00 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm