

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122921\
 Data File : PO083962.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Dec 2021 8:07
 Operator : AJ\MA
 Sample : AR1254ICC050
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC050

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 01/03/2022
 Supervised By :mohammad ahmed 01/03/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 30 08:25:54 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122921.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 30 08:25:44 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.653	3.632	637929	136525	4.852	4.793
2) SA Decachlor...	10.499	8.903	380029	60872	4.922	4.354
Target Compounds						
26) L6 AR-1254-1	6.876	5.750	202539	41331	51.393	49.620
27) L6 AR-1254-2	7.106	5.913	300589	37890	50.843	51.012
28) L6 AR-1254-3	7.489	6.332	252728	63050	45.066	52.677
29) L6 AR-1254-4	7.789	6.579	174597	23298	51.947	51.369m
30) L6 AR-1254-5	8.214	7.003	210438	41438	51.289	43.879m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122921\
Data File : P0083962.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Dec 2021 8:07
Operator : AJ\MA
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC050

Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/03/2022
Supervised By :mohammad ahmed 01/03/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 30 08:25:54 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122921.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 30 08:25:44 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

