

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122922\
 Data File : PO091753.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Dec 2022 21:08
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 29 22:29:02 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 22 03:53:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.289	3.467	116.7E6	51800313	42.951	51.082
2) SA Decachlor...	10.027	8.461	85187750	44539346	45.271	51.052
Target Compounds						
26) L6 AR-1254-1	6.325	5.352	39941395	29921546	402.424	545.115 #
27) L6 AR-1254-2	6.545	5.501	61457059	26228449	410.083	522.966 #
28) L6 AR-1254-3	6.913	5.905	62804824	38198160	419.782	504.053
29) L6 AR-1254-4	7.200	6.135	47125969	23378966	423.860	499.469
30) L6 AR-1254-5	7.622	6.554	48910975	34522529	404.343	492.832



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO122922\
Data File : PO091753.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Dec 2022 21:08
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 29 22:29:02 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO122122.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 22 03:53:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.289 min
Delta R.T.: -0.005 min
Response: 116692588
Conc: 42.95 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.467 min
Delta R.T.: -0.002 min
Response: 51800313
Conc: 51.08 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.027 min
Delta R.T.: -0.004 min
Response: 85187750
Conc: 45.27 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.461 min
Delta R.T.: -0.004 min
Response: 44539346
Conc: 51.05 ng/ml



#26 AR-1254-1


R.T.: 6.325 min
Delta R.T.: -0.003 min
Response: 39941395
Conc: 402.42 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.352 min
Delta R.T.: -0.002 min
Response: 29921546
Conc: 545.12 ng/ml




#27 AR-1254-2

R.T.: 6.545 min
Delta R.T.: -0.004 min
Response: 61457059
Conc: 410.08 ng/ml



#27 AR-1254-2


R.T.: 5.501 min
Delta R.T.: -0.003 min
Response: 26228449
Conc: 522.97 ng/ml



#28 AR-1254-3


R.T.: 6.913 min
Delta R.T.: -0.004 min
Response: 62804824
Conc: 419.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 5.905 min
Delta R.T.: -0.003 min
Response: 38198160
Conc: 504.05 ng/ml




#29 AR-1254-4

R.T.: 7.200 min
Delta R.T.: -0.004 min
Response: 47125969
Conc: 423.86 ng/ml



#29 AR-1254-4


R.T.: 6.135 min
Delta R.T.: -0.002 min
Response: 23378966
Conc: 499.47 ng/ml



#30 AR-1254-5

R.T.: 7.622 min
Delta R.T.: -0.003 min
Response: 48910975
Conc: 404.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.554 min
Delta R.T.: -0.003 min
Response: 34522529
Conc: 492.83 ng/ml