

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0123119\
 Data File : P0065107.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Dec 2019 21:11
 Operator : AJ/MA
 Sample : AR1248ICC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 31 07:27:30 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0123119.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 31 07:26:58 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.179	3.452	1147367	1413263	50.000	50.000
2)	SA Decachlor...	9.697	8.366	1605405	2045375	50.000	50.000
Target Compounds							
21)	L5 AR-1248-1	5.331	4.514	309594	376669	500.000	500.000
22)	L5 AR-1248-2	5.601	4.747	451655	542566	500.000	500.000
23)	L5 AR-1248-3	5.801	4.787	515319	557666	500.000	500.000
24)	L5 AR-1248-4	6.201	4.957	647468	677375	500.000	500.000
25)	L5 AR-1248-5	6.239	5.344	624027	719212	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO123119\
Data File : PO065107.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Dec 2019 21:11
Operator : AJ/MA
Sample : AR1248ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 31 07:27:30 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO123119.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 31 07:26:58 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.179 min
Delta R.T.: 0.000 min
Response: 1147367
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.452 min
Delta R.T.: 0.000 min
Response: 1413263
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.697 min
Delta R.T.: 0.000 min
Response: 1605405
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.366 min
Delta R.T.: 0.000 min
Response: 2045375
Conc: 50.00 ng/ml



#21 AR-1248-1


R.T.: 5.331 min
Delta R.T.: 0.000 min
Response: 309594
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC500




#21 AR-1248-1

R.T.: 4.514 min
Delta R.T.: 0.000 min
Response: 376669
Conc: 500.00 ng/ml




#22 AR-1248-2

R.T.: 5.601 min
Delta R.T.: 0.000 min
Response: 451655
Conc: 500.00 ng/ml



#22 AR-1248-2


R.T.: 4.747 min
Delta R.T.: 0.000 min
Response: 542566
Conc: 500.00 ng/ml



#23 AR-1248-3


R.T.: 5.801 min
Delta R.T.: 0.000 min
Response: 515319
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC500




#23 AR-1248-3

R.T.: 4.787 min
Delta R.T.: 0.000 min
Response: 557666
Conc: 500.00 ng/ml




#24 AR-1248-4

R.T.: 6.201 min
Delta R.T.: 0.000 min
Response: 647468
Conc: 500.00 ng/ml



#24 AR-1248-4


R.T.: 4.957 min
Delta R.T.: 0.000 min
Response: 677375
Conc: 500.00 ng/ml



#25 AR-1248-5

R.T.: 6.239 min
Delta R.T.: 0.000 min
Response: 624027
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC500



#25 AR-1248-5

R.T.: 5.344 min
Delta R.T.: 0.000 min
Response: 719212
Conc: 500.00 ng/ml