

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0123119\
 Data File : P0065157.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 31 Dec 2019 12:05
 Operator : AJ/MA
 Sample : K6359-02
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 CH-WALLS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 31 12:18:32 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0123119.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 31 07:42:34 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.190	3.455	543096	678975	25.749	26.876
2) SA Decachlor...	9.712	8.373	477248	585827	15.639	15.222

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO123119\
Data File : P0065157.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 31 Dec 2019 12:05
Operator : AJ/MA
Sample : K6359-02
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
CH-WALLS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 31 12:18:32 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO123119.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 31 07:42:34 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.190 min
Delta R.T.: 0.011 min
Response: 543096
Conc: 25.75 ng/ml

Instrument :
ECD_O
ClientSampleId :
CH-WALLS




#1 Tetrachloro-m-xylene

R.T.: 3.455 min
Delta R.T.: 0.002 min
Response: 678975
Conc: 26.88 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.712 min
Delta R.T.: 0.013 min
Response: 477248
Conc: 15.64 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.373 min
Delta R.T.: 0.006 min
Response: 585827
Conc: 15.22 ng/ml