

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0123119\
 Data File : P0065165.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 31 Dec 2019 16:06
 Operator : AJ/MA
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 31 16:20:39 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0123119.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 31 07:42:34 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.179	3.452	1150582	1390474	54.551	55.040
2) SA Decachlor...	9.700	8.370	1338415	1688023	43.858	43.862
Target Compounds						
3) L1 AR-1016-1	5.332	4.516	551808	656117	513.979	511.456
4) L1 AR-1016-2	5.353	4.534	809839	928891	520.352	514.243
5) L1 AR-1016-3	5.414	4.707	495137	500374	517.652	496.773
6) L1 AR-1016-4	5.511	4.750	411854	418909	521.743	475.091
7) L1 AR-1016-5	5.802	4.959	471766	549060	572.446	498.514
31) L7 AR-1260-1	6.916	5.980	810645	1115174	485.078	469.872
32) L7 AR-1260-2	7.173	6.168	1019580	1390348	496.261	455.718
33) L7 AR-1260-3	7.528	6.319	773753	1266392	484.838	471.486
34) L7 AR-1260-4	7.751	6.786	896107	1061351	482.172	458.024
35) L7 AR-1260-5	8.058	7.029	1730143	2493596	486.505	461.014



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO123119\
Data File : P0065165.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 31 Dec 2019 16:06
Operator : AJ/MA
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 31 16:20:39 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO123119.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 31 07:42:34 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.179 min
Delta R.T.: 0.000 min
Response: 1150582
Conc: 54.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.452 min
Delta R.T.: 0.000 min
Response: 1390474
Conc: 55.04 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.700 min
Delta R.T.: 0.001 min
Response: 1338415
Conc: 43.86 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.370 min
Delta R.T.: 0.002 min
Response: 1688023
Conc: 43.86 ng/ml



#3 AR-1016-1


R.T.: 5.332 min
Delta R.T.: 0.000 min
Response: 551808
Conc: 513.98 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.516 min
Delta R.T.: 0.000 min
Response: 656117
Conc: 511.46 ng/ml




#4 AR-1016-2

R.T.: 5.353 min
Delta R.T.: 0.000 min
Response: 809839
Conc: 520.35 ng/ml



#4 AR-1016-2


R.T.: 4.534 min
Delta R.T.: 0.001 min
Response: 928891
Conc: 514.24 ng/ml



#5 AR-1016-3


R.T.: 5.414 min
Delta R.T.: 0.000 min
Response: 495137
Conc: 517.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.707 min
Delta R.T.: 0.001 min
Response: 500374
Conc: 496.77 ng/ml




#6 AR-1016-4

R.T.: 5.511 min
Delta R.T.: 0.000 min
Response: 411854
Conc: 521.74 ng/ml



#6 AR-1016-4


R.T.: 4.750 min
Delta R.T.: 0.002 min
Response: 418909
Conc: 475.09 ng/ml



#7 AR-1016-5


R.T.: 5.802 min
Delta R.T.: 0.000 min
Response: 471766
Conc: 572.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 4.959 min
Delta R.T.: 0.001 min
Response: 549060
Conc: 498.51 ng/ml




#31 AR-1260-1

R.T.: 6.916 min
Delta R.T.: 0.000 min
Response: 810645
Conc: 485.08 ng/ml



#31 AR-1260-1


R.T.: 5.980 min
Delta R.T.: 0.002 min
Response: 1115174
Conc: 469.87 ng/ml



#32 AR-1260-2


R.T.: 7.173 min
Delta R.T.: 0.000 min
Response: 1019580
Conc: 496.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.168 min
Delta R.T.: 0.002 min
Response: 1390348
Conc: 455.72 ng/ml




#33 AR-1260-3

R.T.: 7.528 min
Delta R.T.: 0.000 min
Response: 773753
Conc: 484.84 ng/ml



#33 AR-1260-3


R.T.: 6.319 min
Delta R.T.: 0.002 min
Response: 1266392
Conc: 471.49 ng/ml



#34 AR-1260-4


R.T.: 7.751 min
Delta R.T.: 0.000 min
Response: 896107
Conc: 482.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 6.786 min
Delta R.T.: 0.001 min
Response: 1061351
Conc: 458.02 ng/ml



#35 AR-1260-5

R.T.: 8.058 min
Delta R.T.: 0.000 min
Response: 1730143
Conc: 486.51 ng/ml



#35 AR-1260-5

R.T.: 7.029 min
Delta R.T.: 0.002 min
Response: 2493596
Conc: 461.01 ng/ml