

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012925\
 Data File : P0109263.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Jan 2025 04:28
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 30 05:41:01 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 22 03:46:11 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.699	3.696	367.6E6	268.2E6	48.647	50.032
2) SA Decachlor...	8.759	8.711	266.3E6	138.9E6	38.439	40.406
Target Compounds						
21) L5 AR-1248-1	4.795	4.780	78251951	54460304	490.388	507.459
22) L5 AR-1248-2	5.033	5.017	103.0E6	74070730	460.965	481.052
23) L5 AR-1248-3	5.248	5.059	136.1E6	82472795	496.853	498.704
24) L5 AR-1248-4	5.603	5.230	186.9E6	97717411	486.061	510.258
25) L5 AR-1248-5	5.645	5.624	130.8E6	93764330	492.348	511.781



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO012925\
Data File : PO109263.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jan 2025 04:28
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 30 05:41:01 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO012125.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 22 03:46:11 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.699 min
Delta R.T.: 0.000 min
Response: 367597810
Conc: 48.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.696 min
Delta R.T.: -0.023 min
Response: 268190063
Conc: 50.03 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.759 min
Delta R.T.: 0.000 min
Response: 266292649
Conc: 38.44 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.711 min
Delta R.T.: -0.022 min
Response: 138946498
Conc: 40.41 ng/ml



#21 AR-1248-1


R.T.: 4.795 min
Delta R.T.: 0.000 min
Response: 78251951
Conc: 490.39 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.780 min
Delta R.T.: 0.000 min
Response: 54460304
Conc: 507.46 ng/ml




#22 AR-1248-2

R.T.: 5.033 min
Delta R.T.: 0.000 min
Response: 103020799
Conc: 460.97 ng/ml



#22 AR-1248-2


R.T.: 5.017 min
Delta R.T.: 0.000 min
Response: 74070730
Conc: 481.05 ng/ml



#23 AR-1248-3


R.T.: 5.248 min
Delta R.T.: 0.000 min
Response: 136068087
Conc: 496.85 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.059 min
Delta R.T.: 0.000 min
Response: 82472795
Conc: 498.70 ng/ml




#24 AR-1248-4

R.T.: 5.603 min
Delta R.T.: 0.000 min
Response: 186913034
Conc: 486.06 ng/ml



#24 AR-1248-4


R.T.: 5.230 min
Delta R.T.: -0.001 min
Response: 97717411
Conc: 510.26 ng/ml



#25 AR-1248-5

R.T.: 5.645 min
Delta R.T.: 0.000 min
Response: 130801783
Conc: 492.35 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.624 min
Delta R.T.: 0.000 min
Response: 93764330
Conc: 511.78 ng/ml