

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO102418\
 Data File : P0050478.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Oct 2018 15:26
 Operator : SM/SJ
 Sample : J5604-05
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 MS-VB-29736-BOX-2

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Oct 24 15:46:59 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101818.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Oct 19 02:02:26 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.357	3.354	5567539	3984622	18.658	20.268
2) SA Decachlor...	10.020	8.104	2361024	2303502	10.881m	12.051m

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO102418\
Data File : PO050478.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 Oct 2018 15:26
Operator : SM/SJ
Sample : J5604-05
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
MS-VB-29736-BOX-2

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Oct 24 15:46:59 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO101818.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Oct 19 02:02:26 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.357 min
Delta R.T.: 0.008 min
Response: 5567539
Conc: 18.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
MS-VB-29736-BOX-2




#1 Tetrachloro-m-xylene

R.T.: 3.354 min
Delta R.T.: -0.006 min
Response: 3984622
Conc: 20.27 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.020 min
Delta R.T.: -0.061 min
Response: 2361024
Conc: 10.88 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.104 min
Delta R.T.: -0.012 min
Response: 2303502
Conc: 12.05 ng/ml m