

Data Path : P:\HPCHEM1\ECD 0\Data\P0010218\
 Data File : P0041004.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 02 Jan 2018 12:51
 Operator : SM/UA
 Sample : I7077-02MS
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 SB-07-COMPMS

Manual Integrations
 APPROVED

UGO
 1/3/2018 10:22:07 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 03 01:21:16 2018
 Quant Method : \\74.0.250.170\pestpcbsrv\HPCHEM1\ECD_0\methods\P0122017.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Dec 20 03:23:29 2017
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.136	3.374	4597532	7026712	22.172m	22.797m
2)	SA Decachlor...	9.554	8.180	3343894	4137201	16.330	11.682m#
Target Compounds							
3)	L1 AR-1016-1	5.007	4.199	1140624	1528469	194.383	167.611
4)	L1 AR-1016-2	5.349	4.596	1550763	1688557	196.635	165.386
5)	L1 AR-1016-3	5.443	4.639	1155406	1342162	184.960	167.420
6)	L1 AR-1016-4	5.732	4.845	1239844	541318	196.126	48.050m#
7)	L1 AR-1016-5	5.868	4.983	992568	905531	145.906	80.684m#
31)	L7 AR-1260-1	6.834	5.840	2000671	3878224	148.760	150.483
32)	L7 AR-1260-2	7.086	6.173	2988120	3951494	151.825	127.296
33)	L7 AR-1260-3	7.361	6.344	2975993	4986997	143.573	133.932
34)	L7 AR-1260-4	7.960	6.874	4249700	7199735	136.960	104.153
35)	L7 AR-1260-5	8.249	7.211	2729205	8209487	153.983m	180.785

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : P:\HPCHEM1\ECD 0\Data\PO010218\
Data File : PO041004.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 02 Jan 2018 12:51
Operator : SM/UA
Sample : I7077-02MS
Misc :
ALS Vial : 10 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
SB-07-COMPMS


Manual Integrations
APPROVED

UGO
1/3/2018 10:22:07 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 03 01:21:16 2018
Quant Method : \\74.0.250.170\pestpcbsrv\HPCHEM1\ECD_0\methods\PO122017.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Dec 20 03:23:29 2017
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Ini. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.136 min
Delta R.T.: -0.010 min
Response: 4597532
Conc: 22.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-07-COMPMS


Manual Integrations
APPROVED

UGO
1/3/2018 10:22:07 AM




#1 Tetrachloro-m-xylene

R.T.: 3.374 min
Delta R.T.: -0.008 min
Response: 7026712
Conc: 22.80 ng/ml m




#2 Decachlorobiphenyl

R.T.: 9.554 min
Delta R.T.: -0.025 min
Response: 3343894
Conc: 16.33 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.180 min
Delta R.T.: -0.015 min
Response: 4137201
Conc: 11.68 ng/ml m




#3 AR-1016-1

R.T.: 5.007 min
Delta R.T.: -0.011 min
Response: 1140624
Conc: 194.38 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-07-COMPMS


Manual Integrations
APPROVED

UGO
1/3/2018 10:22:07 AM




#3 AR-1016-1

R.T.: 4.199 min
Delta R.T.: -0.009 min
Response: 1528469
Conc: 167.61 ng/ml




#4 AR-1016-2

R.T.: 5.349 min
Delta R.T.: -0.011 min
Response: 1550763
Conc: 196.64 ng/ml



#4 AR-1016-2

R.T.: 4.596 min
Delta R.T.: -0.009 min
Response: 1688557
Conc: 165.39 ng/ml




#5 AR-1016-3

R.T.: 5.443 min
Delta R.T.: -0.011 min
Response: 1155406
Conc: 184.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-07-COMPMS


Manual Integrations
APPROVED

UGO
1/3/2018 10:22:07 AM




#5 AR-1016-3

R.T.: 4.639 min
Delta R.T.: -0.009 min
Response: 1342162
Conc: 167.42 ng/ml




#6 AR-1016-4

R.T.: 5.732 min
Delta R.T.: -0.011 min
Response: 1239844
Conc: 196.13 ng/ml



#6 AR-1016-4

R.T.: 4.845 min
Delta R.T.: -0.007 min
Response: 541318
Conc: 48.05 ng/ml m




#7 AR-1016-5

R.T.: 5.868 min
Delta R.T.: -0.012 min
Response: 992568
Conc: 145.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-07-COMPMS


Manual Integrations
APPROVED

UGO
1/3/2018 10:22:07 AM




#7 AR-1016-5

R.T.: 4.983 min
Delta R.T.: -0.011 min
Response: 905531
Conc: 80.68 ng/ml m




#31 AR-1260-1

R.T.: 6.834 min
Delta R.T.: -0.012 min
Response: 2000671
Conc: 148.76 ng/ml



#31 AR-1260-1

R.T.: 5.840 min
Delta R.T.: -0.011 min
Response: 3878224
Conc: 150.48 ng/ml




#32 AR-1260-2

R.T.: 7.086 min
Delta R.T.: -0.013 min
Response: 2988120
Conc: 151.82 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-07-COMPMS


Manual Integrations
APPROVED

UGO
1/3/2018 10:22:07 AM




#32 AR-1260-2

R.T.: 6.173 min
Delta R.T.: -0.011 min
Response: 3951494
Conc: 127.30 ng/ml




#33 AR-1260-3

R.T.: 7.361 min
Delta R.T.: -0.014 min
Response: 2975993
Conc: 143.57 ng/ml



#33 AR-1260-3

R.T.: 6.344 min
Delta R.T.: -0.011 min
Response: 4986997
Conc: 133.93 ng/ml




#34 AR-1260-4

R.T.: 7.960 min
Delta R.T.: -0.015 min
Response: 4249700
Conc: 136.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-07-COMPMS


Manual Integrations
APPROVED

UGO
1/3/2018 10:22:07 AM




#34 AR-1260-4

R.T.: 6.874 min
Delta R.T.: -0.011 min
Response: 7199735
Conc: 104.15 ng/ml



#35 AR-1260-5

R.T.: 8.249 min
Delta R.T.: -0.015 min
Response: 2729205
Conc: 153.98 ng/ml m



#35 AR-1260-5

R.T.: 7.211 min
Delta R.T.: -0.012 min
Response: 8209487
Conc: 180.78 ng/ml