

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0010219\
 Data File : P0053119.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 02 Jan 2019 18:05
 Operator : SM/SJ
 Sample : J6598-02
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 B3_122818

Manual Integrations
 APPROVED

Sohil
 1/3/2019 8:39:34 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 03 07:29:41 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0010219.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 02 14:46:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.182	3.214	19234095	2912722	14.700	12.980m
2) SA Decachlor...	9.653	7.860	8618056	3179119	10.877	12.367

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0010219\
Data File : P0053119.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 02 Jan 2019 18:05
Operator : SM/SJ
Sample : J6598-02
Misc :
ALS Vial : 10 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
B3_122818


Manual Integrations
APPROVED

Sohil
1/3/2019 8:39:34 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 03 07:29:41 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0010219.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 02 14:46:04 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.182 min
Delta R.T.: 0.000 min
Response: 19234095
Conc: 14.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
B3_122818


Manual Integrations
APPROVED

Sohil
1/3/2019 8:39:34 AM




#1 Tetrachloro-m-xylene

R.T.: 3.214 min
Delta R.T.: 0.000 min
Response: 2912722
Conc: 12.98 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.653 min
Delta R.T.: 0.002 min
Response: 8618056
Conc: 10.88 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.860 min
Delta R.T.: 0.002 min
Response: 3179119
Conc: 12.37 ng/ml