

Data Path : P:\HPCHEM1\ECD_0\Data\P0010618\
 Data File : P0041124.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 05 Jan 2018 18:53
 Operator : SM/UA
 Sample : AR1660 MDL5
 Misc : 25PPB SOIL ECD_0
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 AR1660 MDL5

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 06 00:03:26 2018
 Quant Method : \\74.0.250.170\pestpcbsrv\HPCHEM1\ECD_0\methods\P0122917CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 29 01:32:03 2017
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.132	3.373	3718873	5580268	22.578	23.537
2) SA Decachlor...	9.547	8.180	10090689	15767866	44.372	44.033
Target Compounds						
3) L1 AR-1016-1	5.002	4.198	184262	260636	46.010	43.357
4) L1 AR-1016-2	5.344	4.595	272104	380492	48.090	53.791
5) L1 AR-1016-3	5.439	4.638	189444	332406	42.629	54.506 #
6) L1 AR-1016-4	5.726	4.676	227271	388641	48.453	54.672
7) L1 AR-1016-5	5.864	4.841	275652	404173	55.815	52.108
31) L7 AR-1260-1	6.829	5.839	441436	956065	43.225	51.282
32) L7 AR-1260-2	7.082	6.025	686138	1256999	45.581	46.249
33) L7 AR-1260-3	7.357	6.171	711365	878422	40.000	39.548
34) L7 AR-1260-4	7.656	6.632	516679	785853	46.514	47.907
35) L7 AR-1260-5	7.957	6.873	1153629	2117375	47.426	44.087



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : P:\HPCHEM1\ECD_0\Data\PO010618\
Data File : PO041124.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 05 Jan 2018 18:53
Operator : SM/UA
Sample : AR1660 MDL5
Misc : 25PPB SOIL ECD_0
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
AR1660 MDL5

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 06 00:03:26 2018
Quant Method : \\74.0.250.170\pestpcbsrv\HPCHEM1\ECD_0\methods\PO122917CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 29 01:32:03 2017
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#3 AR-1016-1


R.T.: 5.002 min
Delta R.T.: -0.004 min
Response: 184262
Conc: 46.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660 MDL5




#3 AR-1016-1

R.T.: 4.198 min
Delta R.T.: -0.001 min
Response: 260636
Conc: 43.36 ng/ml




#4 AR-1016-2

R.T.: 5.344 min
Delta R.T.: -0.004 min
Response: 272104
Conc: 48.09 ng/ml



#4 AR-1016-2


R.T.: 4.595 min
Delta R.T.: -0.002 min
Response: 380492
Conc: 53.79 ng/ml



#5 AR-1016-3


R.T.: 5.439 min
Delta R.T.: -0.004 min
Response: 189444
Conc: 42.63 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660 MDL5




#5 AR-1016-3

R.T.: 4.638 min
Delta R.T.: 0.000 min
Response: 332406
Conc: 54.51 ng/ml




#6 AR-1016-4

R.T.: 5.726 min
Delta R.T.: -0.005 min
Response: 227271
Conc: 48.45 ng/ml



#6 AR-1016-4


R.T.: 4.676 min
Delta R.T.: -0.002 min
Response: 388641
Conc: 54.67 ng/ml



#7 AR-1016-5


R.T.: 5.864 min
Delta R.T.: -0.005 min
Response: 275652
Conc: 55.81 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660 MDL5




#7 AR-1016-5

R.T.: 4.841 min
Delta R.T.: -0.002 min
Response: 404173
Conc: 52.11 ng/ml




#31 AR-1260-1

R.T.: 6.829 min
Delta R.T.: -0.004 min
Response: 441436
Conc: 43.23 ng/ml



#31 AR-1260-1


R.T.: 5.839 min
Delta R.T.: -0.002 min
Response: 956065
Conc: 51.28 ng/ml



#32 AR-1260-2


R.T.: 7.082 min
Delta R.T.: -0.004 min
Response: 686138
Conc: 45.58 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660 MDL5




#32 AR-1260-2

R.T.: 6.025 min
Delta R.T.: 0.000 min
Response: 1256999
Conc: 46.25 ng/ml




#33 AR-1260-3

R.T.: 7.357 min
Delta R.T.: -0.003 min
Response: 711365
Conc: 40.00 ng/ml



#33 AR-1260-3


R.T.: 6.171 min
Delta R.T.: -0.001 min
Response: 878422
Conc: 39.55 ng/ml



#34 AR-1260-4


R.T.: 7.656 min
Delta R.T.: -0.003 min
Response: 516679
Conc: 46.51 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660 MDL5




#34 AR-1260-4

R.T.: 6.632 min
Delta R.T.: 0.000 min
Response: 785853
Conc: 47.91 ng/ml



#35 AR-1260-5

R.T.: 7.957 min
Delta R.T.: -0.004 min
Response: 1153629
Conc: 47.43 ng/ml



#35 AR-1260-5

R.T.: 6.873 min
Delta R.T.: -0.001 min
Response: 2117375
Conc: 44.09 ng/ml