

Data Path : P:\HPCHEM1\ECD_0\Data\P0010818\
Data File : P0041217.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 08 Jan 2018 11:03
Operator : SM/UA
Sample : J1004-08
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
SB-16-COMP

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 09 01:24:10 2018
Quant Method : P:\HPCHEM1\ECD_0\methods\P0122017.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Dec 20 03:23:29 2017
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.137	3.370	3159720	5032120	15.238	16.326m
2) SA Decachlor...	9.551	8.177	1660376	2446691	8.108	6.909m

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : P:\HPCHEM1\ECD_0\Data\P0010818\
Data File : P0041217.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 08 Jan 2018 11:03
Operator : SM/UA
Sample : J1004-08
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
SB-16-COMP

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 09 01:24:10 2018
Quant Method : P:\HPCHEM1\ECD_0\methods\P0122017.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Dec 20 03:23:29 2017
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.137 min
Delta R.T.: -0.009 min
Response: 3159720
Conc: 15.24 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-16-COMP




#1 Tetrachloro-m-xylene

R.T.: 3.370 min
Delta R.T.: -0.013 min
Response: 5032120
Conc: 16.33 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.551 min
Delta R.T.: -0.028 min
Response: 1660376
Conc: 8.11 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.177 min
Delta R.T.: -0.018 min
Response: 2446691
Conc: 6.91 ng/ml m