

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0010819\
 Data File : P0053317.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 08 Jan 2019 16:18
 Operator : SM/SJ
 Sample : K1074-01
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 09 03:41:48 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0010819.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 09 03:34:05 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.172	3.208	26189188	4477761	19.599	17.908
2)	SA Decachlor...	9.634	7.853	9908381	3083992	12.545	13.467
Target Compounds							
6)	L1 AR-1016-4	0.000	4.455	0	767876	N.D.	127.145 #
7)	L1 AR-1016-5	5.777	0.000	2524279	0	106.909	N.D. #
14)	L3 AR-1232-4	0.000	4.455	0	767876	N.D.	226.692 #
19)	L4 AR-1242-4	0.000	4.455	0	767876	N.D.	113.798 #
22)	L5 AR-1248-2	0.000	4.455	0	767876	N.D.	85.685 #
23)	L5 AR-1248-3	5.777	4.455	2524279	767876	73.714	83.077
27)	L6 AR-1254-2	6.424	5.112	1458658	736065	22.785	44.851 #
32)	L7 AR-1260-2	7.143	0.000	379351	0	7.920	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO010819\
Data File : PO053317.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 08 Jan 2019 16:18
Operator : SM/SJ
Sample : K1074-01
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 09 03:41:48 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO010819.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 09 03:34:05 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.172 min
Delta R.T.: 0.000 min
Response: 26189188
Conc: 19.60 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.208 min
Delta R.T.: 0.000 min
Response: 4477761
Conc: 17.91 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.634 min
Delta R.T.: 0.000 min
Response: 9908381
Conc: 12.54 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.853 min
Delta R.T.: 0.000 min
Response: 3083992
Conc: 13.47 ng/ml



#6 AR-1016-4


R.T.: 0.000 min
Exp R.T.: 5.498 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#6 AR-1016-4

R.T.: 4.455 min
Delta R.T.: 0.043 min
Response: 767876
Conc: 127.15 ng/ml




#7 AR-1016-5

R.T.: 5.777 min
Delta R.T.: -0.007 min
Response: 2524279
Conc: 106.91 ng/ml



#7 AR-1016-5


R.T.: 0.000 min
Exp R.T.: 4.605 min
Response: 0
Conc: N.D.



#14 AR-1232-4


R.T.: 0.000 min
Exp R.T. : 5.498 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#14 AR-1232-4

R.T.: 4.455 min
Delta R.T.: 0.006 min
Response: 767876
Conc: 226.69 ng/ml




#19 AR-1242-4

R.T.: 0.000 min
Exp R.T. : 5.498 min
Response: 0
Conc: N.D.



#19 AR-1242-4


R.T.: 4.455 min
Delta R.T.: 0.006 min
Response: 767876
Conc: 113.80 ng/ml



#22 AR-1248-2


R.T.: 0.000 min
Exp R.T.: 5.584 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#22 AR-1248-2

R.T.: 4.455 min
Delta R.T.: 0.042 min
Response: 767876
Conc: 85.69 ng/ml




#23 AR-1248-3

R.T.: 5.777 min
Delta R.T.: -0.007 min
Response: 2524279
Conc: 73.71 ng/ml



#23 AR-1248-3


R.T.: 4.455 min
Delta R.T.: 0.007 min
Response: 767876
Conc: 83.08 ng/ml



#27 AR-1254-2


R.T.: 6.424 min
Delta R.T.: 0.056 min
Response: 1458658
Conc: 22.79 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 5.112 min
Delta R.T.: 0.041 min
Response: 736065
Conc: 44.85 ng/ml



#32 AR-1260-2

R.T.: 7.143 min
Delta R.T.: 0.000 min
Response: 379351
Conc: 7.92 ng/ml



#32 AR-1260-2

R.T.: 0.000 min
Exp R.T. : 5.750 min
Response: 0
Conc: N.D.