

Data Path : P:\HPCHEM1\ECD_0\Data\P0011418\
 Data File : P0041391.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 12 Jan 2018 9:28
 Operator : SM/UA
 Sample : J1013-01
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 NB-08-010918

Manual Integrations
 APPROVED

ugo
 1/12/2018 7:20:24 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 12 10:13:28 2018
 Quant Method : P:\HPCHEM1\ECD_0\methods\P0122017.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Dec 20 03:23:29 2017
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.138	3.372	4955919	7134957	23.900	23.148m
2) SA Decachlor...	9.555	8.179	3851277	5303546	18.807	14.976m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : P:\HPCHEM1\ECD_0\Data\P0011418\
Data File : P0041391.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 12 Jan 2018 9:28
Operator : SM/UA
Sample : J1013-01
Misc :
ALS Vial : 4 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
NB-08-010918


Manual Integrations
APPROVED

ugo
1/12/2018 7:20:24 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 12 10:13:28 2018
Quant Method : P:\HPCHEM1\ECD_0\methods\P0122017.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Dec 20 03:23:29 2017
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.138 min
Delta R.T.: -0.008 min
Response: 4955919
Conc: 23.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
NB-08-010918


Manual Integrations
APPROVED

ugo
1/12/2018 7:20:24 PM




#1 Tetrachloro-m-xylene

R.T.: 3.372 min
Delta R.T.: -0.011 min
Response: 7134957
Conc: 23.15 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.555 min
Delta R.T.: -0.024 min
Response: 3851277
Conc: 18.81 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.179 min
Delta R.T.: -0.016 min
Response: 5303546
Conc: 14.98 ng/ml m