

Method Path : P:\HPCHEM1\ECD_0\methods\
 Method File : AR1262P0011918CLP.M
 Title : GC EXTRACTABLES
 Last Update : Thu Jan 18 01:30:25 2018
 Response Via : Initial Calibration

Calibration Files

400 =P0041517.D 1600 =P0041519.D 800 =P0041518.D
 200 =P0041516.D 100 =P0041515.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.488	1.383	1.411	1.595	1.724	1.520	E5	9.24
2) SA	Decachlorobiphenyl	1.878	1.639	1.627	2.020	2.237	1.880	E5	13.81
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1	7.480	6.680	7.212	8.091	8.689	7.630	E3	10.22
37) L8	AR-1262-2	1.047	0.898	0.959	1.172	1.106	1.037	E4	10.65
38) L8	AR-1262-3	1.623	1.429	1.523	1.688	1.866	1.626	E4	10.23
39) L8	AR-1262-4	1.273	1.134	1.209	1.338	1.479	1.287	E4	10.23
40) L8	AR-1262-5	2.736	2.507	2.623	2.404	3.193	2.693	E4	11.37
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =P0041517.D 1600 =P0041519.D 800 =P0041518.D
 200 =P0041516.D 100 =P0041515.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.272	2.105	2.155	2.389	2.614	2.307	E5	8.83
2) SA	Decachlorobiphenyl	2.863	2.702	2.780	3.055	3.168	2.913	E5	6.64

Method Path : P:\HPCHEM1\ECD_0\methods\
 Method File : AR1262P0011918CLP.M
 Title : GC EXTRACTABLES
 Last Update : Thu Jan 18 01:30:25 2018
 Response Via : Initial Calibration

Calibration Files

400 =P0041517.D 1600 =P0041519.D 800 =P0041518.D
 200 =P0041516.D 100 =P0041515.D

	Compound	400	1600	800	200	100	Avg	%RSD
3)	L1 AR-1016-1						0.000	-1.00
4)	L1 AR-1016-2						0.000	-1.00
5)	L1 AR-1016-3						0.000	-1.00
6)	L1 AR-1016-4						0.000	-1.00
7)	L1 AR-1016-5						0.000	-1.00
8)	L2 AR-1221-1						0.000	-1.00
9)	L2 AR-1221-2						0.000	-1.00
10)	L2 AR-1221-3						0.000	-1.00
11)	L3 AR-1232-1						0.000	-1.00
12)	L3 AR-1232-2						0.000	-1.00
13)	L3 AR-1232-3						0.000	-1.00
14)	L3 AR-1232-4						0.000	-1.00
15)	L3 AR-1232-5						0.000	-1.00
16)	L4 AR-1242-1						0.000	-1.00
17)	L4 AR-1242-2						0.000	-1.00
18)	L4 AR-1242-3						0.000	-1.00
19)	L4 AR-1242-4						0.000	-1.00
20)	L4 AR-1242-5						0.000	-1.00
21)	L5 AR-1248-1						0.000	-1.00
22)	L5 AR-1248-2						0.000	-1.00
23)	L5 AR-1248-3						0.000	-1.00
24)	L5 AR-1248-4						0.000	-1.00
25)	L5 AR-1248-5						0.000	-1.00
26)	L6 AR-1254-1						0.000	-1.00
27)	L6 AR-1254-2						0.000	-1.00
28)	L6 AR-1254-3						0.000	-1.00
29)	L6 AR-1254-4						0.000	-1.00
30)	L6 AR-1254-5						0.000	-1.00
31)	L7 AR-1260-1						0.000	-1.00
32)	L7 AR-1260-2						0.000	-1.00
33)	L7 AR-1260-3						0.000	-1.00
34)	L7 AR-1260-4						0.000	-1.00
35)	L7 AR-1260-5						0.000	-1.00
36)	L8 AR-1262-1	1.359	1.187	1.258	1.579	1.515	1.380 E4	12.04
37)	L8 AR-1262-2	1.800	1.637	1.701	1.978	2.067	1.837 E4	9.92
38)	L8 AR-1262-3	2.967	2.802	2.885	3.183	3.376	3.043 E4	7.70
39)	L8 AR-1262-4	4.992	4.832	4.893	5.230	5.551	5.099 E4	5.77
40)	L8 AR-1262-5	1.534	1.487	1.512	1.560	1.700	1.558 E4	5.37
41)	L9 AR-1268-1						0.000	-1.00
42)	L9 AR-1268-2						0.000	-1.00
43)	L9 AR-1268-3						0.000	-1.00
44)	L9 AR-1268-4						0.000	-1.00
45)	L9 AR-1268-5						0.000	-1.00

(#) = Out of Range