

Method Path : P:\HPCHEM1\ECD_0\
 Method File : P0012318.M
 Title : GC EXTRACTABLES
 Last Update : Tue Jan 23 00:46:54 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0041665.D 1000 =P0041663.D 750 =P0041664.D
 250 =P0041666.D 50 =P0041667.D

Compound			500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene		1.916	1.778	1.870	1.895	2.192	1.930	E5	8.07
2) SA	Decachlorobiphenyl		2.302	2.066	2.198	2.347	2.834	2.350	E5	12.41
3) L1	AR-1016-1		5.159	4.541	4.920	5.155	6.113	5.178	E3	11.21
4) L1	AR-1016-2		7.321	6.446	6.954	7.430	7.712	7.172	E3	6.81
5) L1	AR-1016-3		6.025	5.382	5.814	5.942	6.632	5.959	E3	7.56
6) L1	AR-1016-4		5.952	5.252	5.736	6.108	7.213	6.052	E3	11.98
7) L1	AR-1016-5		6.420	5.699	6.224	6.408	7.546	6.460	E3	10.44
8) L2	AR-1221-1							0.000		-1.00
9) L2	AR-1221-2							0.000		-1.00
10) L2	AR-1221-3							0.000		-1.00
11) L3	AR-1232-1							0.000		-1.00
12) L3	AR-1232-2							0.000		-1.00
13) L3	AR-1232-3							0.000		-1.00
14) L3	AR-1232-4							0.000		-1.00
15) L3	AR-1232-5							0.000		-1.00
16) L4	AR-1242-1							0.000		-1.00
17) L4	AR-1242-2							0.000		-1.00
18) L4	AR-1242-3							0.000		-1.00
19) L4	AR-1242-4							0.000		-1.00
20) L4	AR-1242-5							0.000		-1.00
21) L5	AR-1248-1							0.000		-1.00
22) L5	AR-1248-2							0.000		-1.00
23) L5	AR-1248-3							0.000		-1.00
24) L5	AR-1248-4							0.000		-1.00
25) L5	AR-1248-5							0.000		-1.00
26) L6	AR-1254-1							0.000		-1.00
27) L6	AR-1254-2							0.000		-1.00
28) L6	AR-1254-3							0.000		-1.00
29) L6	AR-1254-4							0.000		-1.00
30) L6	AR-1254-5							0.000		-1.00
31) L7	AR-1260-1		1.245	1.088	1.202	1.268	1.487	1.258	E4	11.56
32) L7	AR-1260-2		1.728	1.518	1.661	1.750	2.041	1.740	E4	11.00
33) L7	AR-1260-3		2.012	1.788	1.921	2.021	2.294	2.007	E4	9.25
34) L7	AR-1260-4		1.180	1.047	1.123	1.197	1.334	1.176	E4	9.00
35) L7	AR-1260-5		2.628	2.408	2.549	2.608	2.958	2.630	E4	7.69
36) L8	AR-1262-1							0.000		-1.00
37) L8	AR-1262-2							0.000		-1.00
38) L8	AR-1262-3							0.000		-1.00
39) L8	AR-1262-4							0.000		-1.00
40) L8	AR-1262-5							0.000		-1.00
41) L9	AR-1268-1							0.000		-1.00
42) L9	AR-1268-2							0.000		-1.00
43) L9	AR-1268-3							0.000		-1.00
44) L9	AR-1268-4							0.000		-1.00
45) L9	AR-1268-5							0.000		-1.00

Signal #2 Calibration Files

500 =P0041665.D 1000 =P0041663.D 750 =P0041664.D
 250 =P0041666.D 50 =P0041667.D

Compound			500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene		2.944	2.725	2.876	2.863	3.107	2.903	E5	4.79
2) SA	Decachlorobiphenyl		3.558	3.210	3.397	3.546	4.170	3.576	E5	10.08

Method Path : P:\HPCHEM1\ECD_0\
 Method File : P0012318.M
 Title : GC EXTRACTABLES
 Last Update : Tue Jan 23 00:46:54 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0041665.D 1000 =P0041663.D 750 =P0041664.D
 250 =P0041666.D 50 =P0041667.D

Compound			500	1000	750	250	50	Avg		%RSD

3)	L1	AR-1016-1	7.545	6.715	7.290	7.452	8.576	7.516	E3	8.98
4)	L1	AR-1016-2	7.188	6.369	6.877	7.114	7.564	7.022	E3	6.28
5)	L1	AR-1016-3	8.541	7.566	8.189	8.426	9.616	8.468	E3	8.79
6)	L1	AR-1016-4	0.937	0.831	0.907	0.935	1.049	0.932	E4	8.42
7)	L1	AR-1016-5	1.020	0.913	1.001	0.996	1.026	0.991	E4	4.59
8)	L2	AR-1221-1						0.000		-1.00
9)	L2	AR-1221-2						0.000		-1.00
10)	L2	AR-1221-3						0.000		-1.00
11)	L3	AR-1232-1						0.000		-1.00
12)	L3	AR-1232-2						0.000		-1.00
13)	L3	AR-1232-3						0.000		-1.00
14)	L3	AR-1232-4						0.000		-1.00
15)	L3	AR-1232-5						0.000		-1.00
16)	L4	AR-1242-1						0.000		-1.00
17)	L4	AR-1242-2						0.000		-1.00
18)	L4	AR-1242-3						0.000		-1.00
19)	L4	AR-1242-4						0.000		-1.00
20)	L4	AR-1242-5						0.000		-1.00
21)	L5	AR-1248-1						0.000		-1.00
22)	L5	AR-1248-2						0.000		-1.00
23)	L5	AR-1248-3						0.000		-1.00
24)	L5	AR-1248-4						0.000		-1.00
25)	L5	AR-1248-5						0.000		-1.00
26)	L6	AR-1254-1						0.000		-1.00
27)	L6	AR-1254-2						0.000		-1.00
28)	L6	AR-1254-3						0.000		-1.00
29)	L6	AR-1254-4						0.000		-1.00
30)	L6	AR-1254-5						0.000		-1.00
31)	L7	AR-1260-1	2.169	1.886	2.082	2.118	2.417	2.135	E4	8.94
32)	L7	AR-1260-2	3.135	2.743	3.018	3.130	3.768	3.159	E4	11.90
33)	L7	AR-1260-3	2.653	2.336	2.553	2.654	2.855	2.610	E4	7.22
34)	L7	AR-1260-4	3.436	3.057	3.285	3.401	3.869	3.410	E4	8.69
35)	L7	AR-1260-5	5.458	4.856	5.293	5.371	5.849	5.365	E4	6.64
36)	L8	AR-1262-1						0.000		-1.00
37)	L8	AR-1262-2						0.000		-1.00
38)	L8	AR-1262-3						0.000		-1.00
39)	L8	AR-1262-4						0.000		-1.00
40)	L8	AR-1262-5						0.000		-1.00
41)	L9	AR-1268-1						0.000		-1.00
42)	L9	AR-1268-2						0.000		-1.00
43)	L9	AR-1268-3						0.000		-1.00
44)	L9	AR-1268-4						0.000		-1.00
45)	L9	AR-1268-5						0.000		-1.00

(#) = Out of Range