

Data Path : P:\HPCHEM1\ECD 0\Data\P0012418\  
 Data File : P0041738.D  
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
 Acq On : 23 Jan 2018 13:51  
 Operator : SM/UA  
 Sample : J1210-07RE  
 Misc :  
 ALS Vial : 9 Sample Multiplier: 1

Instrument :  
 ECD\_O  
 ClientSampleId :  
 R-4220RE

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Jan 23 15:30:42 2018  
 Quant Method : P:\HPCHEM1\ECD\_0\P0012318.M  
 Quant Title : GC EXTRACTABLES  
 QLast Update : Tue Jan 23 01:50:31 2018  
 Response via : Initial Calibration  
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl  
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

|       | Compound                    | RT#1  | RT#2  | Resp#1  | Resp#2  | ng/ml   | ng/ml   |
|-------|-----------------------------|-------|-------|---------|---------|---------|---------|
| ----- |                             |       |       |         |         |         |         |
|       | System Monitoring Compounds |       |       |         |         |         |         |
| 1)    | SA Tetrachlo...             | 4.123 | 3.365 | 2873095 | 4441061 | 14.886m | 15.299m |
| 2)    | SA Decachlor...             | 9.534 | 8.166 | 2417678 | 3068300 | 10.290m | 8.580m  |

Target Compounds

-----



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : P:\HPCHEM1\ECD 0\Data\P0012418\  
Data File : P0041738.D  
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
Acq On : 23 Jan 2018 13:51  
Operator : SM/UA  
Sample : J1210-07RE  
Misc :  
ALS Vial : 9 Sample Multiplier: 1

Instrument :  
ECD\_O  
ClientSampleId :  
R-4220RE

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Jan 23 15:30:42 2018  
Quant Method : P:\HPCHEM1\ECD\_0\P0012318.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Tue Jan 23 01:50:31 2018  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Ini. : 2 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.123 min  
Delta R.T.: 0.000 min  
Response: 2873095  
Conc: 14.89 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
R-4220RE




#1 Tetrachloro-m-xylene

R.T.: 3.365 min  
Delta R.T.: -0.001 min  
Response: 4441061  
Conc: 15.30 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.534 min  
Delta R.T.: 0.006 min  
Response: 2417678  
Conc: 10.29 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.166 min  
Delta R.T.: 0.004 min  
Response: 3068300  
Conc: 8.58 ng/ml m