

Data Path : P:\HPCHEM1\ECD 0\Data\P0012818\
 Data File : P0041847.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 26 Jan 2018 18:20
 Operator : SM/UA
 Sample : PB106040BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB106040BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 27 02:21:49 2018
 Quant Method : P:\HPCHEM1\ECD_0\P0012318.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 23 01:50:31 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

	System Monitoring Compounds						
1)	SA Tetrachlo...	4.128	3.362	3223606	4739875	16.702	16.328
2)	SA Decachlor...	9.530	8.161	3804722	4836714	16.193	13.525

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : P:\HPCHEM1\ECD 0\Data\P0012818\
Data File : P0041847.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 26 Jan 2018 18:20
Operator : SM/UA
Sample : PB106040BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB106040BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 27 02:21:49 2018
Quant Method : P:\HPCHEM1\ECD_0\P0012318.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 23 01:50:31 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Ini. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.128 min
Delta R.T.: 0.005 min
Response: 3223606
Conc: 16.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB106040BL




#1 Tetrachloro-m-xylene

R.T.: 3.362 min
Delta R.T.: -0.003 min
Response: 4739875
Conc: 16.33 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.530 min
Delta R.T.: 0.001 min
Response: 3804722
Conc: 16.19 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.161 min
Delta R.T.: 0.000 min
Response: 4836714
Conc: 13.52 ng/ml