

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0012819\
 Data File : P0053577.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 Jan 2019 14:01
 Operator : SM/SJ
 Sample : K1234-03DL 2X
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 EWPDL

Manual Integrations
 APPROVED

Sohil
 1/29/2019 11:29:36 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 29 04:40:52 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0012519.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jan 28 05:58:35 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.127	3.172	19697318	2250319	15.202	9.484 #
2) SA Decachlor...	9.537	7.792	14925820	4065424	17.053	16.254
Target Compounds						
26) L6 AR-1254-1	6.102	4.883	12212511	4197258	330.012	251.408m
27) L6 AR-1254-2	6.315	5.021	19631492	3151719	340.515	224.153m#
28) L6 AR-1254-3	6.678	5.394	28302844	5232986	466.689	235.825m#
29) L6 AR-1254-4	6.957	5.614	11602039	3231443	248.226	218.030m
30) L6 AR-1254-5	7.368	6.006	14054143	27412294	305.846	1420.582 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012819\
Data File : P0053577.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 28 Jan 2019 14:01
Operator : SM/SJ
Sample : K1234-03DL 2X
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
EWPDL

Manual Integrations
APPROVED

Sohil
1/29/2019 11:29:36 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 29 04:40:52 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012519.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jan 28 05:58:35 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

