

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012819\
 Data File : P0053591.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 Jan 2019 17:58
 Operator : SM/SJ
 Sample : K1250-02
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 DRUM-6

Manual Integrations
 APPROVED

Sohil
 1/29/2019 11:29:41 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 29 04:45:46 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012519.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jan 28 05:58:35 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.112	3.171	25672613	4538761	19.813	19.129
2) SA Decachlor...	9.514	7.787	10634076	2535537	12.150m	10.138

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO012819\
Data File : PO053591.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 28 Jan 2019 17:58
Operator : SM/SJ
Sample : K1250-02
Misc :
ALS Vial : 19 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
DRUM-6


Manual Integrations
APPROVED

Sohil
1/29/2019 11:29:41 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 29 04:45:46 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO012519.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jan 28 05:58:35 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.112 min
Delta R.T.: -0.004 min
Response: 25672613
Conc: 19.81 ng/ml

Instrument :
ECD_O
ClientSampleId :
DRUM-6


Manual Integrations
APPROVED

Sohil
1/29/2019 11:29:41 AM




#1 Tetrachloro-m-xylene

R.T.: 3.171 min
Delta R.T.: 0.000 min
Response: 4538761
Conc: 19.13 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.514 min
Delta R.T.: -0.010 min
Response: 10634076
Conc: 12.15 ng/ml m



#2 Decachlorobiphenyl

R.T.: 7.787 min
Delta R.T.: -0.003 min
Response: 2535537
Conc: 10.14 ng/ml