

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012820\
 Data File : P0065996.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 Jan 2020 11:57
 Operator : DD\AJ
 Sample : L1242-14
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 28 15:25:24 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012320.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 24 03:54:38 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.150	3.426	607848	769058	26.214	28.954
2)	SA Decachlor...	9.647	8.326	595410	579543	14.303	11.926
Target Compounds							
3)	L1 AR-1016-1	5.301	4.485	197306	226824	179.656	177.060
4)	L1 AR-1016-2	5.322	4.502	378644	429950	228.946	228.687
5)	L1 AR-1016-3	5.384	4.675	67128	118498	65.914	122.118 #
6)	L1 AR-1016-4	5.480	4.717	64048	201341	76.381	245.274 #
7)	L1 AR-1016-5	5.770	4.927	134956	227820	158.000	211.308 #
8)	L2 AR-1221-1	4.332	0.000	18260	0	63.167	N.D. #
9)	L2 AR-1221-2	4.447	3.718	15356	59954	69.614	235.934 #
10)	L2 AR-1221-3	4.519	3.792	49443	48613	68.918	59.071
11)	L3 AR-1232-1	4.519	3.792	49443	48613	91.042	76.373
12)	L3 AR-1232-2	5.036	4.502	118890	429950	411.642	559.370 #
13)	L3 AR-1232-3	5.322	4.675	378644	118498	562.141	307.971 #
14)	L3 AR-1232-4	5.480	4.757	64048	205695	190.592	584.138 #
15)	L3 AR-1232-5	5.570	4.927	121464	227820	489.684	571.726
16)	L4 AR-1242-1	5.301	4.485	197306	226824	244.827	244.370
17)	L4 AR-1242-2	5.322	4.502	378644	429950	316.714	310.866
18)	L4 AR-1242-3	5.384	4.675	67128	118498	89.519	161.012 #
19)	L4 AR-1242-4	5.480	4.757	64048	205695	104.699	276.281 #
20)	L4 AR-1242-5	6.208	5.273	301557	344920	362.156	311.574
21)	L5 AR-1248-1	5.301	4.485	197306	226824	319.501	317.933
22)	L5 AR-1248-2	5.570	4.717	121464	201341	137.954	198.638 #
23)	L5 AR-1248-3	5.770	4.757	134956	205695	130.792	200.120 #
24)	L5 AR-1248-4	6.170	4.927	184599	227820	137.006	178.530 #
25)	L5 AR-1248-5	6.208	5.313	301557	267969	233.323	191.992
26)	L6 AR-1254-1	6.141	5.273	238335	344920	179.563	154.256
27)	L6 AR-1254-2	6.360	5.419	359877	231549	165.202	112.732 #
28)	L6 AR-1254-3	6.722	5.816	367110	369246	150.967	108.888 #
29)	L6 AR-1254-4	7.004	6.043	554135	235416	264.410	102.775 #
30)	L6 AR-1254-5	7.419	6.455	408520	290915	190.398	89.960 #
31)	L7 AR-1260-1	6.882	5.944	242612	300340	125.216	123.991
32)	L7 AR-1260-2	7.138	6.133	301570	303286	115.001	95.715
33)	L7 AR-1260-3	7.494	6.282	169132	106256	77.825	38.426 #
34)	L7 AR-1260-4	7.715	6.749	181980	76362	82.096	30.941 #
35)	L7 AR-1260-5	8.024	6.992	438717	293132	90.607	48.220 #
36)	L8 AR-1262-1	7.494	6.547	169132	95770	56.914	69.207
37)	L8 AR-1262-2	8.024	6.992	438717	293132	85.206	46.362 #
38)	L8 AR-1262-3	8.320	7.271	169595	74430	48.308	29.063 #
39)	L8 AR-1262-4	8.419	7.335	778984	239712	284.933	50.974 #
40)	L8 AR-1262-5	8.983	7.828	74548	83521	39.531	39.720
41)	L9 AR-1268-1	8.320	7.271	169595	74430	25.872	9.554 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012820\
 Data File : P0065996.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 Jan 2020 11:57
 Operator : DD\AJ
 Sample : L1242-14
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 28 15:25:24 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012320.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 24 03:54:38 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.419	7.335	778984	239712	132.037	33.098 #
43) L9	AR-1268-3	8.591	7.537	118550	150914	23.527	25.211
44) L9	AR-1268-4	8.983	7.828	74548	83521	35.183	34.237
45) L9	AR-1268-5	9.349	8.097	133864	127037	8.932	7.032



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0012820\
Data File : P0065996.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 28 Jan 2020 11:57
Operator : DD\AJ
Sample : L1242-14
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 28 15:25:24 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012320.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 24 03:54:38 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.150 min
Delta R.T.: -0.005 min
Response: 607848
Conc: 26.21 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.426 min
Delta R.T.: -0.005 min
Response: 769058
Conc: 28.95 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.647 min
Delta R.T.: -0.002 min
Response: 595410
Conc: 14.30 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.326 min
Delta R.T.: -0.002 min
Response: 579543
Conc: 11.93 ng/ml



#3 AR-1016-1


R.T.: 5.301 min
Delta R.T.: -0.003 min
Response: 197306
Conc: 179.66 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 4.485 min
Delta R.T.: -0.004 min
Response: 226824
Conc: 177.06 ng/ml




#4 AR-1016-2

R.T.: 5.322 min
Delta R.T.: -0.003 min
Response: 378644
Conc: 228.95 ng/ml



#4 AR-1016-2


R.T.: 4.502 min
Delta R.T.: -0.005 min
Response: 429950
Conc: 228.69 ng/ml



#5 AR-1016-3


R.T.: 5.384 min
Delta R.T.: -0.002 min
Response: 67128
Conc: 65.91 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 4.675 min
Delta R.T.: -0.004 min
Response: 118498
Conc: 122.12 ng/ml




#6 AR-1016-4

R.T.: 5.480 min
Delta R.T.: -0.004 min
Response: 64048
Conc: 76.38 ng/ml



#6 AR-1016-4


R.T.: 4.717 min
Delta R.T.: -0.004 min
Response: 201341
Conc: 245.27 ng/ml



#7 AR-1016-5


R.T.: 5.770 min
Delta R.T.: -0.003 min
Response: 134956
Conc: 158.00 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 4.927 min
Delta R.T.: -0.003 min
Response: 227820
Conc: 211.31 ng/ml




#8 AR-1221-1

R.T.: 4.332 min
Delta R.T.: -0.031 min
Response: 18260
Conc: 63.17 ng/ml



#8 AR-1221-1


R.T.: 0.000 min
Exp R.T. : 3.640 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 4.447 min
Delta R.T.: 0.000 min
Response: 15356
Conc: 69.61 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 3.718 min
Delta R.T.: -0.005 min
Response: 59954
Conc: 235.93 ng/ml




#10 AR-1221-3

R.T.: 4.519 min
Delta R.T.: -0.004 min
Response: 49443
Conc: 68.92 ng/ml



#10 AR-1221-3


R.T.: 3.792 min
Delta R.T.: -0.005 min
Response: 48613
Conc: 59.07 ng/ml



#11 AR-1232-1


R.T.: 4.519 min
Delta R.T.: -0.003 min
Response: 49443
Conc: 91.04 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 3.792 min
Delta R.T.: -0.005 min
Response: 48613
Conc: 76.37 ng/ml




#12 AR-1232-2

R.T.: 5.036 min
Delta R.T.: -0.003 min
Response: 118890
Conc: 411.64 ng/ml



#12 AR-1232-2


R.T.: 4.502 min
Delta R.T.: -0.005 min
Response: 429950
Conc: 559.37 ng/ml



#13 AR-1232-3


R.T.: 5.322 min
Delta R.T.: -0.003 min
Response: 378644
Conc: 562.14 ng/ml

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 4.675 min
Delta R.T.: -0.004 min
Response: 118498
Conc: 307.97 ng/ml




#14 AR-1232-4

R.T.: 5.480 min
Delta R.T.: -0.004 min
Response: 64048
Conc: 190.59 ng/ml



#14 AR-1232-4


R.T.: 4.757 min
Delta R.T.: -0.004 min
Response: 205695
Conc: 584.14 ng/ml



#15 AR-1232-5


R.T.: 5.570 min
Delta R.T.: -0.003 min
Response: 121464
Conc: 489.68 ng/ml

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 4.927 min
Delta R.T.: -0.003 min
Response: 227820
Conc: 571.73 ng/ml




#16 AR-1242-1

R.T.: 5.301 min
Delta R.T.: -0.004 min
Response: 197306
Conc: 244.83 ng/ml



#16 AR-1242-1


R.T.: 4.485 min
Delta R.T.: -0.005 min
Response: 226824
Conc: 244.37 ng/ml



#17 AR-1242-2


R.T.: 5.322 min
Delta R.T.: -0.003 min
Response: 378644
Conc: 316.71 ng/ml

Instrument :
ECD_O
ClientSampleId :




#17 AR-1242-2

R.T.: 4.502 min
Delta R.T.: -0.005 min
Response: 429950
Conc: 310.87 ng/ml




#18 AR-1242-3

R.T.: 5.384 min
Delta R.T.: -0.002 min
Response: 67128
Conc: 89.52 ng/ml



#18 AR-1242-3


R.T.: 4.675 min
Delta R.T.: -0.005 min
Response: 118498
Conc: 161.01 ng/ml



#19 AR-1242-4


R.T.: 5.480 min
Delta R.T.: -0.005 min
Response: 64048
Conc: 104.70 ng/ml

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 4.757 min
Delta R.T.: -0.005 min
Response: 205695
Conc: 276.28 ng/ml




#20 AR-1242-5

R.T.: 6.208 min
Delta R.T.: -0.002 min
Response: 301557
Conc: 362.16 ng/ml



#20 AR-1242-5


R.T.: 5.273 min
Delta R.T.: -0.004 min
Response: 344920
Conc: 311.57 ng/ml



#21 AR-1248-1


R.T.: 5.301 min
Delta R.T.: -0.003 min
Response: 197306
Conc: 319.50 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 4.485 min
Delta R.T.: -0.004 min
Response: 226824
Conc: 317.93 ng/ml




#22 AR-1248-2

R.T.: 5.570 min
Delta R.T.: -0.003 min
Response: 121464
Conc: 137.95 ng/ml



#22 AR-1248-2


R.T.: 4.717 min
Delta R.T.: -0.004 min
Response: 201341
Conc: 198.64 ng/ml



#23 AR-1248-3


R.T.: 5.770 min
Delta R.T.: -0.003 min
Response: 134956
Conc: 130.79 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 4.757 min
Delta R.T.: -0.005 min
Response: 205695
Conc: 200.12 ng/ml




#24 AR-1248-4

R.T.: 6.170 min
Delta R.T.: -0.003 min
Response: 184599
Conc: 137.01 ng/ml



#24 AR-1248-4


R.T.: 4.927 min
Delta R.T.: -0.003 min
Response: 227820
Conc: 178.53 ng/ml



#25 AR-1248-5


R.T.: 6.208 min
Delta R.T.: -0.003 min
Response: 301557
Conc: 233.32 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 5.313 min
Delta R.T.: -0.004 min
Response: 267969
Conc: 191.99 ng/ml




#26 AR-1254-1

R.T.: 6.141 min
Delta R.T.: -0.004 min
Response: 238335
Conc: 179.56 ng/ml



#26 AR-1254-1


R.T.: 5.273 min
Delta R.T.: -0.003 min
Response: 344920
Conc: 154.26 ng/ml



#27 AR-1254-2


R.T.: 6.360 min
Delta R.T.: -0.003 min
Response: 359877
Conc: 165.20 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 5.419 min
Delta R.T.: -0.003 min
Response: 231549
Conc: 112.73 ng/ml




#28 AR-1254-3

R.T.: 6.722 min
Delta R.T.: -0.002 min
Response: 367110
Conc: 150.97 ng/ml



#28 AR-1254-3


R.T.: 5.816 min
Delta R.T.: -0.004 min
Response: 369246
Conc: 108.89 ng/ml



#29 AR-1254-4


R.T.: 7.004 min
Delta R.T.: -0.003 min
Response: 554135
Conc: 264.41 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 6.043 min
Delta R.T.: -0.003 min
Response: 235416
Conc: 102.78 ng/ml




#30 AR-1254-5

R.T.: 7.419 min
Delta R.T.: -0.003 min
Response: 408520
Conc: 190.40 ng/ml



#30 AR-1254-5


R.T.: 6.455 min
Delta R.T.: -0.004 min
Response: 290915
Conc: 89.96 ng/ml



#31 AR-1260-1


R.T.: 6.882 min
Delta R.T.: -0.003 min
Response: 242612
Conc: 125.22 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 5.944 min
Delta R.T.: -0.004 min
Response: 300340
Conc: 123.99 ng/ml




#32 AR-1260-2

R.T.: 7.138 min
Delta R.T.: -0.003 min
Response: 301570
Conc: 115.00 ng/ml



#32 AR-1260-2


R.T.: 6.133 min
Delta R.T.: -0.003 min
Response: 303286
Conc: 95.71 ng/ml



#33 AR-1260-3


R.T.: 7.494 min
Delta R.T.: -0.003 min
Response: 169132
Conc: 77.83 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 6.282 min
Delta R.T.: -0.005 min
Response: 106256
Conc: 38.43 ng/ml




#34 AR-1260-4

R.T.: 7.715 min
Delta R.T.: -0.005 min
Response: 181980
Conc: 82.10 ng/ml



#34 AR-1260-4


R.T.: 6.749 min
Delta R.T.: -0.004 min
Response: 76362
Conc: 30.94 ng/ml



#35 AR-1260-5


R.T.: 8.024 min
Delta R.T.: 0.000 min
Response: 438717
Conc: 90.61 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 6.992 min
Delta R.T.: -0.003 min
Response: 293132
Conc: 48.22 ng/ml




#36 AR-1262-1

R.T.: 7.494 min
Delta R.T.: -0.003 min
Response: 169132
Conc: 56.91 ng/ml



#36 AR-1262-1


R.T.: 6.547 min
Delta R.T.: -0.002 min
Response: 95770
Conc: 69.21 ng/ml



#37 AR-1262-2


R.T.: 8.024 min
Delta R.T.: 0.000 min
Response: 438717
Conc: 85.21 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 6.992 min
Delta R.T.: -0.003 min
Response: 293132
Conc: 46.36 ng/ml




#38 AR-1262-3

R.T.: 8.320 min
Delta R.T.: 0.012 min
Response: 169595
Conc: 48.31 ng/ml



#38 AR-1262-3


R.T.: 7.271 min
Delta R.T.: -0.002 min
Response: 74430
Conc: 29.06 ng/ml



#39 AR-1262-4


R.T.: 8.419 min
Delta R.T.: 0.028 min
Response: 778984
Conc: 284.93 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 7.335 min
Delta R.T.: -0.004 min
Response: 239712
Conc: 50.97 ng/ml




#40 AR-1262-5

R.T.: 8.983 min
Delta R.T.: -0.002 min
Response: 74548
Conc: 39.53 ng/ml



#40 AR-1262-5


R.T.: 7.828 min
Delta R.T.: -0.003 min
Response: 83521
Conc: 39.72 ng/ml



#41 AR-1268-1


R.T.: 8.320 min
Delta R.T.: 0.013 min
Response: 169595
Conc: 25.87 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 7.271 min
Delta R.T.: -0.002 min
Response: 74430
Conc: 9.55 ng/ml




#42 AR-1268-2

R.T.: 8.419 min
Delta R.T.: 0.026 min
Response: 778984
Conc: 132.04 ng/ml



#42 AR-1268-2


R.T.: 7.335 min
Delta R.T.: -0.004 min
Response: 239712
Conc: 33.10 ng/ml



#43 AR-1268-3


R.T.: 8.591 min
Delta R.T.: -0.006 min
Response: 118550
Conc: 23.53 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 7.537 min
Delta R.T.: -0.003 min
Response: 150914
Conc: 25.21 ng/ml




#44 AR-1268-4

R.T.: 8.983 min
Delta R.T.: -0.002 min
Response: 74548
Conc: 35.18 ng/ml



#44 AR-1268-4


R.T.: 7.828 min
Delta R.T.: -0.002 min
Response: 83521
Conc: 34.24 ng/ml



#45 AR-1268-5

R.T.: 9.349 min
Delta R.T.: -0.004 min
Response: 133864
Conc: 8.93 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.097 min
Delta R.T.: -0.003 min
Response: 127037
Conc: 7.03 ng/ml