

Data Path : P:\HPCHEM1\ECD_0\Data\P0012918\
 Data File : P0041857.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 Jan 2018 8:58
 Operator : SM/UA
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 29 00:47:18 2018
 Quant Method : P:\HPCHEM1\ECD_0\P0012318.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 23 01:50:31 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.127	3.363	5525439	8329695	28.628	28.694
2) SA Decachlor...	9.529	8.162	6108441	8730106	25.998	24.412
Target Compounds						
12) L3 AR-1232-2	5.200	0.000	329310	0	125.132	N.D. #
32) L7 AR-1260-2	0.000	6.009	0	508918	N.D.	16.112 #
35) L7 AR-1260-5	7.944	6.859	315472	571608	11.994	10.654
36) L8 AR-1262-1	0.000	6.009	0	508918	N.D.	22.913 #
40) L8 AR-1262-5	7.944	6.859	315472	571608	9.800	9.022
41) L9 AR-1268-1	0.000	7.196	0	395884	N.D.	5.402 #
42) L9 AR-1268-2	0.000	7.196	0	395884	N.D.	6.123 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : P:\HPCHEM1\ECD_0\Data\P0012918\
Data File : P0041857.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 28 Jan 2018 8:58
Operator : SM/UA
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 29 00:47:18 2018
Quant Method : P:\HPCHEM1\ECD_0\P0012318.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 23 01:50:31 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.127 min
Delta R.T.: 0.004 min
Response: 5525439
Conc: 28.63 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.363 min
Delta R.T.: -0.003 min
Response: 8329695
Conc: 28.69 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.529 min
Delta R.T.: 0.000 min
Response: 6108441
Conc: 26.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.162 min
Delta R.T.: 0.000 min
Response: 8730106
Conc: 24.41 ng/ml



#12 AR-1232-2


R.T.: 5.200 min
Delta R.T.: -0.051 min
Response: 329310
Conc: 125.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#12 AR-1232-2

R.T.: 0.000 min
Exp R.T.: 4.415 min
Response: 0
Conc: N.D.




#32 AR-1260-2

R.T.: 0.000 min
Exp R.T.: 7.068 min
Response: 0
Conc: N.D.



#32 AR-1260-2


R.T.: 6.009 min
Delta R.T.: -0.002 min
Response: 508918
Conc: 16.11 ng/ml



#35 AR-1260-5


R.T.: 7.944 min
Delta R.T.: 0.001 min
Response: 315472
Conc: 11.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#35 AR-1260-5

R.T.: 6.859 min
Delta R.T.: 0.000 min
Response: 571608
Conc: 10.65 ng/ml




#36 AR-1262-1

R.T.: 0.000 min
Exp R.T.: 6.814 min
Response: 0
Conc: N.D.



#36 AR-1262-1


R.T.: 6.009 min
Delta R.T.: -0.002 min
Response: 508918
Conc: 22.91 ng/ml



#40 AR-1262-5


R.T.: 7.944 min
Delta R.T.: 0.003 min
Response: 315472
Conc: 9.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#40 AR-1262-5

R.T.: 6.859 min
Delta R.T.: 0.002 min
Response: 571608
Conc: 9.02 ng/ml




#41 AR-1268-1

R.T.: 0.000 min
Exp R.T.: 8.221 min
Response: 0
Conc: N.D.



#41 AR-1268-1


R.T.: 7.196 min
Delta R.T.: 0.063 min
Response: 395884
Conc: 5.40 ng/ml



#42 AR-1268-2

R.T.: 0.000 min
Exp R.T.: 8.305 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
I.BLK



#42 AR-1268-2

R.T.: 7.196 min
Delta R.T.: -0.002 min
Response: 395884
Conc: 6.12 ng/ml