

Data Path : P:\HPCHEM1\ECD 0\Data\P0012918\
 Data File : P0041861.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 Jan 2018 10:02
 Operator : SM/UA
 Sample : J1349-02 10X
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 B307522

Manual Integrations
 APPROVED

ugo
 1/29/2018 2:35:20 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 29 04:40:27 2018
 Quant Method : P:\HPCHEM1\ECD_0\P0012318.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 23 01:50:31 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.122	3.366	341838	355495	1.771	1.225 #
2) SA Decachlor...	9.521	8.161	349545	386636	1.488m	1.081 #

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : P:\HPCHEM1\ECD 0\Data\PO012918\
Data File : PO041861.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 28 Jan 2018 10:02
Operator : SM/UA
Sample : J1349-02 10X
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
B307522


Manual Integrations
APPROVED

ugo
1/29/2018 2:35:20 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 29 04:40:27 2018
Quant Method : P:\HPCHEM1\ECD_0\PO012318.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 23 01:50:31 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Ini. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.122 min
Delta R.T.: -0.002 min
Response: 341838
Conc: 1.77 ng/ml

Instrument :
ECD_O
ClientSampleId :
B307522


Manual Integrations
APPROVED

ugo
1/29/2018 2:35:20 PM




#1 Tetrachloro-m-xylene

R.T.: 3.366 min
Delta R.T.: 0.000 min
Response: 355495
Conc: 1.22 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.521 min
Delta R.T.: -0.007 min
Response: 349545
Conc: 1.49 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.161 min
Delta R.T.: -0.001 min
Response: 386636
Conc: 1.08 ng/ml