

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0020119\
 Data File : P0053657.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 01 Feb 2019 19:09
 Operator : SM/SJ
 Sample : K1307-01
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 013019-SIDI-E-U-B

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 01 21:44:37 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0013019.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 01 08:56:03 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.098	3.161	28143117	4023801	14.439	16.776
2) SA Decachlor...	9.487	7.768	13292393	3730344	13.638	14.939
Target Compounds						
8) L2 AR-1221-1	0.000	3.310	0	187027	N.D.	74.729 #
10) L2 AR-1221-3	4.491	3.568	3353843	486107	85.151	77.194
11) L3 AR-1232-1	4.491	3.568	3353843	486107	104.431	90.779
31) L7 AR-1260-1	0.000	5.493	0	431613	N.D.	28.345 #
33) L7 AR-1260-3	0.000	5.887	0	337336	N.D.	19.733 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0020119\
Data File : P0053657.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 01 Feb 2019 19:09
Operator : SM/SJ
Sample : K1307-01
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
013019-SIDI-E-U-B

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 01 21:44:37 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0013019.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 01 08:56:03 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.098 min
Delta R.T.: -0.006 min
Response: 28143117
Conc: 14.44 ng/ml

Instrument :
ECD_O
ClientSampleId :
013019-SIDI-E-U-B




#1 Tetrachloro-m-xylene

R.T.: 3.161 min
Delta R.T.: -0.004 min
Response: 4023801
Conc: 16.78 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.487 min
Delta R.T.: -0.014 min
Response: 13292393
Conc: 13.64 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.768 min
Delta R.T.: -0.008 min
Response: 3730344
Conc: 14.94 ng/ml



#8 AR-1221-1


R.T.: 0.000 min
Exp R.T.: 4.315 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
013019-SIDI-E-U-B




#8 AR-1221-1

R.T.: 3.310 min
Delta R.T.: -0.048 min
Response: 187027
Conc: 74.73 ng/ml




#10 AR-1221-3

R.T.: 4.491 min
Delta R.T.: 0.018 min
Response: 3353843
Conc: 85.15 ng/ml



#10 AR-1221-3


R.T.: 3.568 min
Delta R.T.: 0.066 min
Response: 486107
Conc: 77.19 ng/ml



#11 AR-1232-1


R.T.: 4.491 min
Delta R.T.: 0.018 min
Response: 3353843
Conc: 104.43 ng/ml

Instrument :
ECD_O
ClientSampleId :
013019-SIDI-E-U-B




#11 AR-1232-1

R.T.: 3.568 min
Delta R.T.: 0.066 min
Response: 486107
Conc: 90.78 ng/ml




#31 AR-1260-1

R.T.: 0.000 min
Exp R.T. : 6.806 min
Response: 0
Conc: N.D.



#31 AR-1260-1


R.T.: 5.493 min
Delta R.T.: -0.009 min
Response: 431613
Conc: 28.35 ng/ml



#33 AR-1260-3

R.T.: 0.000 min
Exp R.T.: 7.409 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
013019-SIDI-E-U-B



#33 AR-1260-3

R.T.: 5.887 min
Delta R.T.: 0.064 min
Response: 337336
Conc: 19.73 ng/ml