

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0020121\
 Data File : P0075391.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 01 Feb 2021 15:25
 Operator : AJ/MA
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 01 16:34:29 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0012821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 29 09:09:46 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.925	3.996	5498854	2607540	54.335	54.243
2) SA Decachlor...	10.944	9.247	5704665	2794715	59.127	57.574
Target Compounds						
26) L6 AR-1254-1	7.162	6.104	2277684	1584965	528.480	511.745
27) L6 AR-1254-2	7.392	6.263	3548631	1374348	526.335	511.797
28) L6 AR-1254-3	7.776	6.680	3663191	2197065	531.197	515.259
29) L6 AR-1254-4	8.073	6.918	3203678	1502241	529.926	515.402
30) L6 AR-1254-5	8.503	7.343	3113037	2027869	556.906	522.495



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO020121\
Data File : PO075391.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 01 Feb 2021 15:25
Operator : AJ/MA
Sample : AR1254CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 01 16:34:29 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO012821.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 29 09:09:46 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.925 min
Delta R.T.: 0.001 min
Response: 5498854
Conc: 54.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.996 min
Delta R.T.: 0.000 min
Response: 2607540
Conc: 54.24 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.944 min
Delta R.T.: 0.006 min
Response: 5704665
Conc: 59.13 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.247 min
Delta R.T.: 0.000 min
Response: 2794715
Conc: 57.57 ng/ml



#26 AR-1254-1


R.T.: 7.162 min
Delta R.T.: 0.002 min
Response: 2277684
Conc: 528.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 6.104 min
Delta R.T.: 0.000 min
Response: 1584965
Conc: 511.74 ng/ml




#27 AR-1254-2

R.T.: 7.392 min
Delta R.T.: 0.002 min
Response: 3548631
Conc: 526.33 ng/ml



#27 AR-1254-2


R.T.: 6.263 min
Delta R.T.: 0.000 min
Response: 1374348
Conc: 511.80 ng/ml



#28 AR-1254-3


R.T.: 7.776 min
Delta R.T.: 0.003 min
Response: 3663191
Conc: 531.20 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.680 min
Delta R.T.: 0.000 min
Response: 2197065
Conc: 515.26 ng/ml




#29 AR-1254-4

R.T.: 8.073 min
Delta R.T.: 0.002 min
Response: 3203678
Conc: 529.93 ng/ml



#29 AR-1254-4


R.T.: 6.918 min
Delta R.T.: 0.000 min
Response: 1502241
Conc: 515.40 ng/ml



#30 AR-1254-5

R.T.: 8.503 min
Delta R.T.: 0.003 min
Response: 3113037
Conc: 556.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 7.343 min
Delta R.T.: 0.000 min
Response: 2027869
Conc: 522.49 ng/ml