

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0020419\
 Data File : P0053690.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 04 Feb 2019 11:12
 Operator : SM/SJ
 Sample : K1303-05RE
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 131-U-XRE

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 04 23:36:16 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0013019.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 01 08:56:03 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.107	3.160	19174114	2254842	9.838	9.401
2) SA Decachlor...	9.499	7.771	7986685	2046723	8.195	8.197

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0020419\
Data File : P0053690.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 04 Feb 2019 11:12
Operator : SM/SJ
Sample : K1303-05RE
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
131-U-XRE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 04 23:36:16 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0013019.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 01 08:56:03 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.107 min
Delta R.T.: 0.003 min
Response: 19174114
Conc: 9.84 ng/ml

Instrument :
ECD_O
ClientSampleId :
131-U-XRE




#1 Tetrachloro-m-xylene

R.T.: 3.160 min
Delta R.T.: -0.004 min
Response: 2254842
Conc: 9.40 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.499 min
Delta R.T.: -0.002 min
Response: 7986685
Conc: 8.19 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.771 min
Delta R.T.: -0.006 min
Response: 2046723
Conc: 8.20 ng/ml