

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0020620\
 Data File : P0066328.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 06 Feb 2020 12:09
 Operator : DD\AJ
 Sample : L1379-06
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 M030-2C

Manual Integrations
 APPROVED

mohammad
 2/7/2020 5:04:58 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 07 03:30:02 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0013020.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 31 01:14:46 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.149	3.427	1015096	1176796	34.846	28.769
2) SA Decachlor...	9.644	8.324	698789	959933	17.732	16.378
Target Compounds						
26) L6 AR-1254-1	6.146	5.279	184155	382149	118.526	120.055
27) L6 AR-1254-2	6.359	5.418	451854	90928	181.844	30.983 #
28) L6 AR-1254-3	6.730	5.816	539175	164761	201.779	35.878 #
29) L6 AR-1254-4	7.003	6.042	183040	171178	79.269m	54.327 #
30) L6 AR-1254-5	7.419	6.454	339555	999572	148.813	236.001 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0020620\
Data File : P0066328.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 06 Feb 2020 12:09
Operator : DD\AJ
Sample : L1379-06
Misc :
ALS Vial : 13 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
M030-2C


Manual Integrations
APPROVED

mohammad
2/7/2020 5:04:58 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 07 03:30:02 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0013020.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 31 01:14:46 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.149 min
Delta R.T.: -0.016 min
Response: 1015096
Conc: 34.85 ng/ml

Instrument :
ECD_O
ClientSampleId :
M030-2C


Manual Integrations
APPROVED

mohammad
2/7/2020 5:04:58 PM




#1 Tetrachloro-m-xylene

R.T.: 3.427 min
Delta R.T.: -0.006 min
Response: 1176796
Conc: 28.77 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.644 min
Delta R.T.: -0.018 min
Response: 698789
Conc: 17.73 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.324 min
Delta R.T.: -0.008 min
Response: 959933
Conc: 16.38 ng/ml




#26 AR-1254-1

R.T.: 6.146 min
Delta R.T.: 0.001 min
Response: 184155
Conc: 118.53 ng/ml

Instrument :
ECD_O
ClientSampleId :
M030-2C


Manual Integrations
APPROVED

mohammad
2/7/2020 5:04:58 PM




#26 AR-1254-1

R.T.: 5.279 min
Delta R.T.: 0.004 min
Response: 382149
Conc: 120.06 ng/ml




#27 AR-1254-2

R.T.: 6.359 min
Delta R.T.: -0.002 min
Response: 451854
Conc: 181.84 ng/ml



#27 AR-1254-2

R.T.: 5.418 min
Delta R.T.: -0.003 min
Response: 90928
Conc: 30.98 ng/ml




#28 AR-1254-3

R.T.: 6.730 min
Delta R.T.: 0.006 min
Response: 539175
Conc: 201.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
M030-2C


Manual Integrations
APPROVED

mohammad
2/7/2020 5:04:58 PM




#28 AR-1254-3

R.T.: 5.816 min
Delta R.T.: -0.002 min
Response: 164761
Conc: 35.88 ng/ml




#29 AR-1254-4

R.T.: 7.003 min
Delta R.T.: -0.003 min
Response: 183040
Conc: 79.27 ng/ml m



#29 AR-1254-4

R.T.: 6.042 min
Delta R.T.: -0.003 min
Response: 171178
Conc: 54.33 ng/ml




#30 AR-1254-5

R.T.: 7.419 min
Delta R.T.: -0.002 min
Response: 339555
Conc: 148.81 ng/ml

Instrument :
ECD_O
ClientSampleId :
M030-2C

Manual Integrations
APPROVED

mohammad
2/7/2020 5:04:58 PM



#30 AR-1254-5

R.T.: 6.454 min
Delta R.T.: -0.003 min
Response: 999572
Conc: 236.00 ng/ml