

Data Path : P:\HPCHEM1\ECD 0\Data\P0020718\
Data File : P0042143.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 04 Feb 2018 13:59
Operator : SM/UA
Sample : PB106270BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB106270BL

Manual Integrations
APPROVED

ugo
2/5/2018 12:52:57 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 04 21:58:17 2018
Quant Method : P:\HPCHEM1\ECD_0\P0012318.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 23 01:50:31 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	nq/ml	nq/ml

	System Monitoring Compounds						
1)	SA Tetrachlo...	4.121	3.358	4624600	6271380	23.961m	21.604
2)	SA Decachlor...	9.520	8.155	4504220	6896551	19.171	19.285

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : P:\HPCHEM1\ECD 0\Data\P0020718\
Data File : P0042143.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 04 Feb 2018 13:59
Operator : SM/UA
Sample : PB106270BL
Misc :
ALS Vial : 4 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
PB106270BL


Manual Integrations
APPROVED

ugo
2/5/2018 12:52:57 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 04 21:58:17 2018
Quant Method : P:\HPCHEM1\ECD_0\P0012318.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 23 01:50:31 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Ini. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.121 min
Delta R.T.: -0.002 min
Response: 4624600
Conc: 23.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB106270BL


Manual Integrations
APPROVED

ugo
2/5/2018 12:52:57 PM




#1 Tetrachloro-m-xylene

R.T.: 3.358 min
Delta R.T.: -0.007 min
Response: 6271380
Conc: 21.60 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.520 min
Delta R.T.: -0.009 min
Response: 4504220
Conc: 19.17 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.155 min
Delta R.T.: -0.007 min
Response: 6896551
Conc: 19.28 ng/ml