

Method Path : P:\HPCHEM1\ECD_0\
 Method File : P0021418.M
 Title : GC EXTRACTABLES
 Last Update : Wed Feb 14 15:43:23 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0042392.D 1000 =P0042390.D 750 =P0042391.D
 250 =P0042393.D 50 =P0042394.D

Compound			500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene		1.559	1.525	1.505	1.574	1.571	1.547 E5	1.97
2) SA	Decachlorobiphenyl		1.779	1.701	1.680	1.631	1.804	1.719 E5	4.15
3) L1	AR-1016-1		4.717	4.475	4.492	4.930	4.663	4.656 E3	4.00
4) L1	AR-1016-2		6.361	6.095	6.025	6.493	6.071	6.209 E3	3.32
5) L1	AR-1016-3		5.400	5.272	5.191	5.595	5.443	5.380 E3	2.91
6) L1	AR-1016-4		5.407	5.118	5.050	5.618	5.270	5.293 E3	4.32
7) L1	AR-1016-5		5.863	5.553	5.527	6.287	4.850	5.616 E3	9.38
8) L2	AR-1221-1		1.915					1.915 E3	0.00
9) L2	AR-1221-2		1.459					1.459 E3	0.00
10) L2	AR-1221-3		4.559					4.559 E3	0.00
11) L3	AR-1232-1		3.286					3.286 E3	0.00
12) L3	AR-1232-2		2.487					2.487 E3	0.00
13) L3	AR-1232-3		3.987					3.987 E3	0.00
14) L3	AR-1232-4		2.335					2.335 E3	0.00
15) L3	AR-1232-5		1.943					1.943 E3	0.00
16) L4	AR-1242-1		2.872					2.872 E3	0.00
17) L4	AR-1242-2		6.634					6.634 E3	0.00
18) L4	AR-1242-3		3.370					3.370 E3	0.00
19) L4	AR-1242-4		3.411					3.411 E3	0.00
20) L4	AR-1242-5		3.838					3.838 E3	0.00
21) L5	AR-1248-1		6.478					6.478 E3	0.00
22) L5	AR-1248-2		5.647					5.647 E3	0.00
23) L5	AR-1248-3		7.854					7.854 E3	0.00
24) L5	AR-1248-4		7.854					7.854 E3	0.00
25) L5	AR-1248-5		5.349					5.349 E3	0.00
26) L6	AR-1254-1		1.249	1.114	1.267	1.330	1.629	1.318 E4	14.49
27) L6	AR-1254-2		7.558	6.897	7.768	7.790	8.630	7.729 E3	8.02
28) L6	AR-1254-3		1.420	1.260	1.432	1.513	1.922	1.509 E4	16.45
29) L6	AR-1254-4		1.171	1.034	1.173	1.238	1.433	1.210 E4	12.01
30) L6	AR-1254-5		8.252	7.448	8.447	9.056	9.362	8.513 E3	8.76
31) L7	AR-1260-1		1.114	1.068	1.048	1.133	1.192	1.111 E4	5.13
32) L7	AR-1260-2		1.503	1.449	1.415	1.535	1.549	1.490 E4	3.82
33) L7	AR-1260-3		1.691	1.647	1.596	1.690	1.623	1.650 E4	2.53
34) L7	AR-1260-4		1.214	1.180	1.152	1.234	1.158	1.188 E4	2.98
35) L7	AR-1260-5		2.609	2.572	2.485	2.593	2.576	2.567 E4	1.87
36) L8	AR-1262-1		8.227					8.227 E3	0.00
37) L8	AR-1262-2		1.087					1.087 E4	0.00
38) L8	AR-1262-3		9.961					9.961 E3	0.00
39) L8	AR-1262-4		1.339					1.339 E4	0.00
40) L8	AR-1262-5		2.803					2.803 E4	0.00
41) L9	AR-1268-1		3.358					3.358 E4	0.00
42) L9	AR-1268-2		2.968					2.968 E4	0.00
43) L9	AR-1268-3		2.524					2.524 E4	0.00
44) L9	AR-1268-4		1.088					1.088 E4	0.00
45) L9	AR-1268-5		7.160					7.160 E4	0.00

Signal #2 Calibration Files

500 =P0042392.D 1000 =P0042390.D 750 =P0042391.D
 250 =P0042393.D 50 =P0042394.D

Compound			500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene		2.301	2.275	2.231	2.297	2.251	2.271 E5	1.32
2) SA	Decachlorobiphenyl		2.569	2.476	2.457	2.552	2.511	2.513 E5	1.91

Method Path : P:\HPCHEM1\ECD_0\
 Method File : P0021418.M
 Title : GC EXTRACTABLES
 Last Update : Wed Feb 14 15:43:23 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0042392.D 1000 =P0042390.D 750 =P0042391.D
 250 =P0042393.D 50 =P0042394.D

Compound			500	1000	750	250	50	Avg		%RSD

3)	L1	AR-1016-1	7.012	6.739	6.708	6.679	6.507	6.729	E3	2.70
4)	L1	AR-1016-2	5.871	5.507	5.537	6.101	5.558	5.715	E3	4.57
5)	L1	AR-1016-3	7.358	7.001	7.023	7.659	7.552	7.319	E3	4.10
6)	L1	AR-1016-4	8.080	7.726	7.707	8.422	7.704	7.928	E3	4.02
7)	L1	AR-1016-5	8.254	7.900	7.948	8.222	7.262	7.917	E3	5.04
8)	L2	AR-1221-1	2.584					2.584	E3	0.00
9)	L2	AR-1221-2	1.929					1.929	E3	0.00
10)	L2	AR-1221-3	6.204					6.204	E3	0.00
11)	L3	AR-1232-1	4.359					4.359	E3	0.00
12)	L3	AR-1232-2	8.804					8.804	E3	0.00
13)	L3	AR-1232-3	3.415					3.415	E3	0.00
14)	L3	AR-1232-4	2.662					2.662	E3	0.00
15)	L3	AR-1232-5	1.900					1.900	E3	0.00
16)	L4	AR-1242-1	6.375					6.375	E3	0.00
17)	L4	AR-1242-2	1.509					1.509	E4	0.00
18)	L4	AR-1242-3	4.489					4.489	E3	0.00
19)	L4	AR-1242-4	4.585					4.585	E3	0.00
20)	L4	AR-1242-5	4.978					4.978	E3	0.00
21)	L5	AR-1248-1	9.632					9.632	E3	0.00
22)	L5	AR-1248-2	8.761					8.761	E3	0.00
23)	L5	AR-1248-3	1.472					1.472	E4	0.00
24)	L5	AR-1248-4	1.187					1.187	E4	0.00
25)	L5	AR-1248-5	7.405					7.405	E3	0.00
26)	L6	AR-1254-1	1.627	1.471	1.679	1.689	1.890	1.671	E4	8.99
27)	L6	AR-1254-2	1.468	1.323	1.514	1.518	1.688	1.502	E4	8.67
28)	L6	AR-1254-3	2.433	2.207	2.509	2.509	2.604	2.453	E4	6.11
29)	L6	AR-1254-4	1.967	1.768	1.748	2.030	2.406	1.984	E4	13.40
30)	L6	AR-1254-5	1.145	1.036	1.168	1.168	1.275	1.159	E4	7.32
31)	L7	AR-1260-1	1.885	1.800	1.797	1.960	1.919	1.872	E4	3.86
32)	L7	AR-1260-2	2.660	2.507	2.519	2.826	3.645	2.831	E4	16.70
33)	L7	AR-1260-3	2.225	2.150	2.130	2.249	2.277	2.206	E4	2.89
34)	L7	AR-1260-4	2.626	2.547	2.533	2.625	2.689	2.604	E4	2.47
35)	L7	AR-1260-5	5.031	4.911	4.812	4.963	4.764	4.896	E4	2.22
36)	L8	AR-1262-1	1.861					1.861	E4	0.00
37)	L8	AR-1262-2	1.672					1.672	E4	0.00
38)	L8	AR-1262-3	3.047					3.047	E4	0.00
39)	L8	AR-1262-4	2.316					2.316	E4	0.00
40)	L8	AR-1262-5	5.225					5.225	E4	0.00
41)	L9	AR-1268-1	6.277					6.277	E4	0.00
42)	L9	AR-1268-2	5.573					5.573	E4	0.00
43)	L9	AR-1268-3	4.698					4.698	E4	0.00
44)	L9	AR-1268-4	1.835					1.835	E4	0.00
45)	L9	AR-1268-5	3.389					3.389	E2	0.00

(#) = Out of Range