

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0021820\
 Data File : P0066632.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 18 Feb 2020 17:51
 Operator : DD\AJ
 Sample : L1565-03 10X
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 14-A-14-B-14-C

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 19 06:51:56 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0021320.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Feb 13 16:40:54 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.127	3.409	57947	64544	1.693	1.602
2)	SA Decachlor...	9.605	8.295	69446	94557	1.514	1.572
Target Compounds							
38)	L8 AR-1262-3	0.000	7.284	0	252410	N.D.	58.423 #
39)	L8 AR-1262-4	0.000	7.284	0	252410	N.D.	36.332 #
41)	L9 AR-1268-1	0.000	7.284	0	252410	N.D.	21.341 #
42)	L9 AR-1268-2	0.000	7.284	0	252410	N.D.	25.660 #
43)	L9 AR-1268-3	8.529	7.447	128206	867109	22.168	103.652 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0021820\
Data File : P0066632.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 18 Feb 2020 17:51
Operator : DD\AJ
Sample : L1565-03 10X
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
14-A-14-B-14-C

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 19 06:51:56 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0021320.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Feb 13 16:40:54 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.127 min
Delta R.T.: -0.008 min
Response: 57947
Conc: 1.69 ng/ml

Instrument :
ECD_O
ClientSampleId :
14-A-14-B-14-C




#1 Tetrachloro-m-xylene

R.T.: 3.409 min
Delta R.T.: -0.003 min
Response: 64544
Conc: 1.60 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.605 min
Delta R.T.: -0.018 min
Response: 69446
Conc: 1.51 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.295 min
Delta R.T.: -0.008 min
Response: 94557
Conc: 1.57 ng/ml



#38 AR-1262-3


R.T.: 0.000 min
Exp R.T. : 8.284 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
14-A-14-B-14-C




#38 AR-1262-3

R.T.: 7.284 min
Delta R.T.: 0.033 min
Response: 252410
Conc: 58.42 ng/ml




#39 AR-1262-4

R.T.: 0.000 min
Exp R.T. : 8.367 min
Response: 0
Conc: N.D.



#39 AR-1262-4


R.T.: 7.284 min
Delta R.T.: -0.032 min
Response: 252410
Conc: 36.33 ng/ml



#41 AR-1268-1


R.T.: 0.000 min
Exp R.T. : 8.282 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
14-A-14-B-14-C




#41 AR-1268-1

R.T.: 7.284 min
Delta R.T.: 0.034 min
Response: 252410
Conc: 21.34 ng/ml




#42 AR-1268-2

R.T.: 0.000 min
Exp R.T. : 8.368 min
Response: 0
Conc: N.D.



#42 AR-1268-2


R.T.: 7.284 min
Delta R.T.: -0.031 min
Response: 252410
Conc: 25.66 ng/ml



#43 AR-1268-3

R.T.: 8.529 min
Delta R.T.: -0.042 min
Response: 128206
Conc: 22.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
14-A-14-B-14-C



#43 AR-1268-3

R.T.: 7.447 min
Delta R.T.: -0.067 min
Response: 867109
Conc: 103.65 ng/ml