

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0021820\
 Data File : P0066635.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 18 Feb 2020 18:42
 Operator : DD\AJ
 Sample : L1572-02 10X
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 10-A-10-B-10-C

Manual Integrations
 APPROVED

Sohil
 2/19/2020 4:07:15 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 19 06:52:58 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0021320.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Feb 13 16:40:54 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.128	3.408	95670	119661	2.796	2.970m
2) SA Decachlor...	9.605	8.295	129452	174118	2.823m	2.895

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO021820\
Data File : PO066635.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 18 Feb 2020 18:42
Operator : DD\AJ
Sample : L1572-02 10X
Misc :
ALS Vial : 29 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
10-A-10-B-10-C


Manual Integrations
APPROVED

Sohil
2/19/2020 4:07:15 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 19 06:52:58 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO021320.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Feb 13 16:40:54 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.128 min
Delta R.T.: -0.007 min
Response: 95670
Conc: 2.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
10-A-10-B-10-C


Manual Integrations
APPROVED

Sohil
2/19/2020 4:07:15 PM




#1 Tetrachloro-m-xylene

R.T.: 3.408 min
Delta R.T.: -0.005 min
Response: 119661
Conc: 2.97 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.605 min
Delta R.T.: -0.017 min
Response: 129452
Conc: 2.82 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.295 min
Delta R.T.: -0.008 min
Response: 174118
Conc: 2.90 ng/ml