

Method Path : P:\HPCHEM1\ECD_0\
 Method File : P0021918.M
 Title : GC EXTRACTABLES
 Last Update : Mon Feb 19 13:19:13 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0042464.D 1000 =P0042462.D 750 =P0042463.D
 250 =P0042465.D 50 =P0042466.D

Compound			500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene		8.390	8.921	8.606	8.579	7.841	8.468	E5	4.71
2) SA	Decachlorobiphenyl		6.710	6.176	6.735	6.980	6.859	6.692	E5	4.60
3) L1	AR-1016-1		1.963	1.940	1.981	2.061	2.019	1.993	E4	2.40
4) L1	AR-1016-2		2.517	2.556	2.535	2.660	2.584	2.570	E4	2.18
5) L1	AR-1016-3		2.090	2.133	2.108	2.202	2.132	2.133	E4	1.99
6) L1	AR-1016-4		1.881	1.906	1.901	1.993	1.925	1.921	E4	2.24
7) L1	AR-1016-5		2.036	2.058	2.084	2.146	2.139	2.093	E4	2.32
8) L2	AR-1221-1		8.482					8.482	E3	0.00
9) L2	AR-1221-2		6.173					6.173	E3	0.00
10) L2	AR-1221-3		2.008					2.008	E4	0.00
11) L3	AR-1232-1		1.460					1.460	E4	0.00
12) L3	AR-1232-2		9.296					9.296	E3	0.00
13) L3	AR-1232-3		1.536					1.536	E4	0.00
14) L3	AR-1232-4		8.954					8.954	E3	0.00
15) L3	AR-1232-5		7.221					7.221	E3	0.00
16) L4	AR-1242-1		7.200					7.200	E3	0.00
17) L4	AR-1242-2		2.592					2.592	E4	0.00
18) L4	AR-1242-3		1.239					1.239	E4	0.00
19) L4	AR-1242-4		1.135					1.135	E4	0.00
20) L4	AR-1242-5		1.267					1.267	E4	0.00
21) L5	AR-1248-1		2.286					2.286	E4	0.00
22) L5	AR-1248-2		2.000					2.000	E4	0.00
23) L5	AR-1248-3		2.736					2.736	E4	0.00
24) L5	AR-1248-4		2.602					2.602	E4	0.00
25) L5	AR-1248-5		1.736					1.736	E4	0.00
26) L6	AR-1254-1		4.206					4.206	E4	0.00
27) L6	AR-1254-2		2.570					2.570	E4	0.00
28) L6	AR-1254-3		4.667					4.667	E4	0.00
29) L6	AR-1254-4		3.801					3.801	E4	0.00
30) L6	AR-1254-5		2.734					2.734	E4	0.00
31) L7	AR-1260-1		3.636	3.661	3.682	3.788	3.698	3.693	E4	1.57
32) L7	AR-1260-2		4.958	5.037	5.053	5.144	5.053	5.049	E4	1.31
33) L7	AR-1260-3		5.721	5.873	5.836	5.865	5.417	5.742	E4	3.34
34) L7	AR-1260-4		4.046	4.100	4.108	4.160	3.656	4.014	E4	5.09
35) L7	AR-1260-5		0.990	1.014	1.012	0.989	0.914	0.984	E5	4.17
36) L8	AR-1262-1		2.699					2.699	E4	0.00
37) L8	AR-1262-2		3.603					3.603	E4	0.00
38) L8	AR-1262-3		3.316					3.316	E4	0.00
39) L8	AR-1262-4		4.596					4.596	E4	0.00
40) L8	AR-1262-5		1.056					1.056	E5	0.00
41) L9	AR-1268-1		1.325					1.325	E5	0.00
42) L9	AR-1268-2		1.173					1.173	E5	0.00
43) L9	AR-1268-3		1.007					1.007	E5	0.00
44) L9	AR-1268-4		4.384					4.384	E4	0.00
45) L9	AR-1268-5		3.010					3.010	E5	0.00

Signal #2 Calibration Files

500 =P0042464.D 1000 =P0042462.D 750 =P0042463.D
 250 =P0042465.D 50 =P0042466.D

Compound			500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene		1.091	1.129	1.127	1.106	0.989	1.089	E6	5.31
2) SA	Decachlorobiphenyl		8.313	7.822	8.401	8.566	7.802	8.181	E5	4.26

Method Path : P:\HPCHEM1\ECD_0\
 Method File : P0021918.M
 Title : GC EXTRACTABLES
 Last Update : Mon Feb 19 13:19:13 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0042464.D 1000 =P0042462.D 750 =P0042463.D
 250 =P0042465.D 50 =P0042466.D

Compound			500	1000	750	250	50	Avg		%RSD

3)	L1	AR-1016-1	2.423	2.299	2.450	2.509	2.374	2.411	E4	3.29
4)	L1	AR-1016-2	2.039	2.007	2.042	2.176	2.455	2.144	E4	8.66
5)	L1	AR-1016-3	2.538	2.458	2.540	2.670	2.949	2.631	E4	7.34
6)	L1	AR-1016-4	2.650	2.641	2.671	2.786	2.783	2.706	E4	2.67
7)	L1	AR-1016-5	2.645	2.740	2.664	2.727	2.317	2.619	E4	6.62
8)	L2	AR-1221-1	1.094					1.094	E4	0.00
9)	L2	AR-1221-2	8.073					8.073	E3	0.00
10)	L2	AR-1221-3	2.570					2.570	E4	0.00
11)	L3	AR-1232-1	1.817					1.817	E4	0.00
12)	L3	AR-1232-2	3.220					3.220	E4	0.00
13)	L3	AR-1232-3	1.341					1.341	E4	0.00
14)	L3	AR-1232-4	9.406					9.406	E3	0.00
15)	L3	AR-1232-5	6.452					6.452	E3	0.00
16)	L4	AR-1242-1	2.306					2.306	E4	0.00
17)	L4	AR-1242-2	5.479					5.479	E4	0.00
18)	L4	AR-1242-3	1.589					1.589	E4	0.00
19)	L4	AR-1242-4	1.543					1.543	E4	0.00
20)	L4	AR-1242-5	1.690					1.690	E4	0.00
21)	L5	AR-1248-1	3.161					3.161	E4	0.00
22)	L5	AR-1248-2	2.769					2.769	E4	0.00
23)	L5	AR-1248-3	4.606					4.606	E4	0.00
24)	L5	AR-1248-4	3.784					3.784	E4	0.00
25)	L5	AR-1248-5	2.174					2.174	E4	0.00
26)	L6	AR-1254-1	5.285					5.285	E4	0.00
27)	L6	AR-1254-2	4.519					4.519	E4	0.00
28)	L6	AR-1254-3	7.294					7.294	E4	0.00
29)	L6	AR-1254-4	5.009					5.009	E4	0.00
30)	L6	AR-1254-5	3.280					3.280	E4	0.00
31)	L7	AR-1260-1	5.563	5.455	5.621	5.763	5.610	5.602	E4	1.99
32)	L7	AR-1260-2	7.985	7.989	8.034	8.292	8.572	8.174	E4	3.13
33)	L7	AR-1260-3	6.543	6.552	6.628	6.746	6.135	6.521	E4	3.53
34)	L7	AR-1260-4	7.878	7.867	7.991	8.134	7.541	7.882	E4	2.78
35)	L7	AR-1260-5	1.563	1.518	1.589	1.600	1.445	1.543	E5	4.11
36)	L8	AR-1262-1	5.707					5.707	E4	0.00
37)	L8	AR-1262-2	4.992					4.992	E4	0.00
38)	L8	AR-1262-3	9.252					9.252	E4	0.00
39)	L8	AR-1262-4	6.894					6.894	E4	0.00
40)	L8	AR-1262-5	1.652					1.652	E5	0.00
41)	L9	AR-1268-1	2.029					2.029	E5	0.00
42)	L9	AR-1268-2	1.813					1.813	E5	0.00
43)	L9	AR-1268-3	1.526					1.526	E5	0.00
44)	L9	AR-1268-4	5.917					5.917	E4	0.00
45)	L9	AR-1268-5	4.057					4.057	E5	0.00

(#) = Out of Range