

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0021919\
 Data File : P0053934.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 19 Feb 2019 12:51
 Operator : SM/SJ
 Sample : K1518-09
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 CABLE-BRIDGE-FOUNDATION-D-5

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 19 23:39:20 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0021819.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Feb 19 03:40:07 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.069	3.139	54561234	4921221	18.193	18.443
2)	SA Decachlor...	9.421	7.725	15246670	3909309	12.142	13.046
Target Compounds							
3)	L1 AR-1016-1	5.198	0.000	2980497	0	42.549	N.D. #
4)	L1 AR-1016-2	5.198	0.000	2980497	0	28.947	N.D. #
5)	L1 AR-1016-3	5.342	4.298	1409926	1505401	23.117	163.030 #
6)	L1 AR-1016-4	5.412	4.298	7903791	1505401	159.461	214.588 #
7)	L1 AR-1016-5	5.660	4.544	7840744	2273264	172.401	239.966 #
10)	L2 AR-1221-3	0.000	3.513	0	2116440	N.D.	173.304 #
11)	L3 AR-1232-1	0.000	3.513	0	2116440	N.D.	388.690 #
13)	L3 AR-1232-3	5.198	4.298	2980497	1505401	73.779	407.686 #
14)	L3 AR-1232-4	5.412	4.377	7903791	865054	418.230	282.143 #
15)	L3 AR-1232-5	5.510	4.544	2015470	2273264	156.998	645.414 #
16)	L4 AR-1242-1	5.198	0.000	2980497	0	56.322	N.D. #
17)	L4 AR-1242-2	5.198	0.000	2980497	0	37.401	N.D. #
18)	L4 AR-1242-3	5.342	4.298	1409926	1505401	30.514	209.798 #
19)	L4 AR-1242-4	5.412	4.377	7903791	865054	208.068	129.738 #
20)	L4 AR-1242-5	6.097	4.833	2035273	3888527	55.793	436.705 #
21)	L5 AR-1248-1	5.198	0.000	2980497	0	72.919	N.D. #
22)	L5 AR-1248-2	5.510	4.298	2015470	1505401	39.481	159.525 #
23)	L5 AR-1248-3	5.660	4.377	7840744	865054	134.295	89.845 #
24)	L5 AR-1248-4	6.030	4.544	14868910	2273264	237.730	189.020
25)	L5 AR-1248-5	6.097	4.833	2035273	3888527	33.637	312.970 #
26)	L6 AR-1254-1	6.030	4.833	14868910	3888527	94.128	80.191
27)	L6 AR-1254-2	6.242	4.971	22067230	2967164	89.980	72.613
28)	L6 AR-1254-3	6.602	5.345	16094639	8031903	63.459	122.005 #
29)	L6 AR-1254-4	6.903	5.561	9094773	1590298	48.221	36.652
30)	L6 AR-1254-5	7.290	5.948	54807526	13243045	306.011	273.299
31)	L7 AR-1260-1	6.752	5.462	67871019	14205710	969.646	744.317
32)	L7 AR-1260-2	7.009	5.644	35080523	7811874	405.635	334.343
33)	L7 AR-1260-3	7.357	5.778	23880781	8238727	439.614	381.263
34)	L7 AR-1260-4	7.583	6.223	27350363	5396738	476.725	386.748
35)	L7 AR-1260-5	7.887	6.464	55834356	22504197	459.692	663.492 #
36)	L8 AR-1262-1	7.357	5.981	23880781	6410212	253.283	269.258
37)	L8 AR-1262-2	7.887	6.464	55834356	22504197	352.685	561.060 #
38)	L8 AR-1262-3	8.181	6.725	21562793	4302663	649.205	272.066 #
39)	L8 AR-1262-4	8.235	6.789	26567604	11412768	332.470	389.580
40)	L8 AR-1262-5	8.810	7.272	17609379	3982528	328.732	290.540
41)	L9 AR-1268-1	8.154	6.725	16672956	4302663	88.221	94.338
42)	L9 AR-1268-2	8.235	6.789	26567604	11412768	150.041	266.373 #
43)	L9 AR-1268-3	8.425	6.973	4400684	1022112	28.013	28.207
44)	L9 AR-1268-4	8.810	7.272	17609379	3982528	295.143	262.200

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0021919\
Data File : P0053934.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 19 Feb 2019 12:51
Operator : SM/SJ
Sample : K1518-09
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
CABLE-BRIDGE-FOUNDATION-D-5

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 19 23:39:20 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0021819.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 19 03:40:07 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
45)	L9 AR-1268-5	9.151	7.526	22946334	5261049	48.883	48.929



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0021919\
Data File : P0053934.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 19 Feb 2019 12:51
Operator : SM/SJ
Sample : K1518-09
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
CABLE-BRIDGE-FOUNDATION-D-5

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 19 23:39:20 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0021819.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 19 03:40:07 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.069 min
Delta R.T.: 0.002 min
Response: 54561234
Conc: 18.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
CABLE-BRIDGE-FOUNDATION-D-5




#1 Tetrachloro-m-xylene

R.T.: 3.139 min
Delta R.T.: -0.001 min
Response: 4921221
Conc: 18.44 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.421 min
Delta R.T.: -0.003 min
Response: 15246670
Conc: 12.14 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.725 min
Delta R.T.: -0.004 min
Response: 3909309
Conc: 13.05 ng/ml



#3 AR-1016-1


R.T.: 5.198 min
Delta R.T.: -0.006 min
Response: 2980497
Conc: 42.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
CABLE-BRIDGE-FOUNDATION-D-5




#3 AR-1016-1

R.T.: 0.000 min
Exp R.T. : 4.112 min
Response: 0
Conc: N.D.




#4 AR-1016-2

R.T.: 5.198 min
Delta R.T.: -0.027 min
Response: 2980497
Conc: 28.95 ng/ml



#4 AR-1016-2


R.T.: 0.000 min
Exp R.T. : 4.126 min
Response: 0
Conc: N.D.



#5 AR-1016-3


R.T.: 5.342 min
Delta R.T.: 0.058 min
Response: 1409926
Conc: 23.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
CABLE-BRIDGE-FOUNDATION-D-5




#5 AR-1016-3

R.T.: 4.298 min
Delta R.T.: 0.016 min
Response: 1505401
Conc: 163.03 ng/ml




#6 AR-1016-4

R.T.: 5.412 min
Delta R.T.: 0.029 min
Response: 7903791
Conc: 159.46 ng/ml



#6 AR-1016-4


R.T.: 4.298 min
Delta R.T.: -0.027 min
Response: 1505401
Conc: 214.59 ng/ml



#7 AR-1016-5


R.T.: 5.660 min
Delta R.T.: -0.004 min
Response: 7840744
Conc: 172.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
CABLE-BRIDGE-FOUNDATION-D-5




#7 AR-1016-5

R.T.: 4.544 min
Delta R.T.: 0.029 min
Response: 2273264
Conc: 239.97 ng/ml




#10 AR-1221-3

R.T.: 0.000 min
Exp R.T. : 4.437 min
Response: 0
Conc: N.D.



#10 AR-1221-3


R.T.: 3.513 min
Delta R.T.: 0.038 min
Response: 2116440
Conc: 173.30 ng/ml



#11 AR-1232-1


R.T.: 0.000 min
Exp R.T.: 4.437 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
CABLE-BRIDGE-FOUNDATION-D-5




#11 AR-1232-1

R.T.: 3.513 min
Delta R.T.: 0.038 min
Response: 2116440
Conc: 388.69 ng/ml




#13 AR-1232-3

R.T.: 5.198 min
Delta R.T.: -0.026 min
Response: 2980497
Conc: 73.78 ng/ml



#13 AR-1232-3


R.T.: 4.298 min
Delta R.T.: 0.016 min
Response: 1505401
Conc: 407.69 ng/ml



#14 AR-1232-4


R.T.: 5.412 min
Delta R.T.: 0.031 min
Response: 7903791
Conc: 418.23 ng/ml

Instrument :
ECD_O
ClientSampleId :
CABLE-BRIDGE-FOUNDATION-D-5




#14 AR-1232-4

R.T.: 4.377 min
Delta R.T.: 0.017 min
Response: 865054
Conc: 282.14 ng/ml




#15 AR-1232-5

R.T.: 5.510 min
Delta R.T.: 0.046 min
Response: 2015470
Conc: 157.00 ng/ml



#15 AR-1232-5


R.T.: 4.544 min
Delta R.T.: 0.030 min
Response: 2273264
Conc: 645.41 ng/ml



#16 AR-1242-1


R.T.: 5.198 min
Delta R.T.: -0.005 min
Response: 2980497
Conc: 56.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
CABLE-BRIDGE-FOUNDATION-D-5




#16 AR-1242-1

R.T.: 0.000 min
Exp R.T. : 4.112 min
Response: 0
Conc: N.D.




#17 AR-1242-2

R.T.: 5.198 min
Delta R.T.: -0.025 min
Response: 2980497
Conc: 37.40 ng/ml



#17 AR-1242-2


R.T.: 0.000 min
Exp R.T. : 4.126 min
Response: 0
Conc: N.D.



#18 AR-1242-3


R.T.: 5.342 min
Delta R.T.: 0.060 min
Response: 1409926
Conc: 30.51 ng/ml

Instrument :
ECD_O
ClientSampleId :
CABLE-BRIDGE-FOUNDATION-D-5




#18 AR-1242-3

R.T.: 4.298 min
Delta R.T.: 0.016 min
Response: 1505401
Conc: 209.80 ng/ml




#19 AR-1242-4

R.T.: 5.412 min
Delta R.T.: 0.031 min
Response: 7903791
Conc: 208.07 ng/ml



#19 AR-1242-4


R.T.: 4.377 min
Delta R.T.: 0.018 min
Response: 865054
Conc: 129.74 ng/ml



#20 AR-1242-5


R.T.: 6.097 min
Delta R.T.: 0.003 min
Response: 2035273
Conc: 55.79 ng/ml

Instrument :
ECD_O
ClientSampleId :
CABLE-BRIDGE-FOUNDATION-D-5




#20 AR-1242-5

R.T.: 4.833 min
Delta R.T.: -0.005 min
Response: 3888527
Conc: 436.70 ng/ml




#21 AR-1248-1

R.T.: 5.198 min
Delta R.T.: -0.004 min
Response: 2980497
Conc: 72.92 ng/ml



#21 AR-1248-1


R.T.: 0.000 min
Exp R.T.: 4.111 min
Response: 0
Conc: N.D.



#22 AR-1248-2


R.T.: 5.510 min
Delta R.T.: 0.047 min
Response: 2015470
Conc: 39.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
CABLE-BRIDGE-FOUNDATION-D-5




#22 AR-1248-2

R.T.: 4.298 min
Delta R.T.: -0.026 min
Response: 1505401
Conc: 159.53 ng/ml




#23 AR-1248-3

R.T.: 5.660 min
Delta R.T.: -0.003 min
Response: 7840744
Conc: 134.29 ng/ml



#23 AR-1248-3


R.T.: 4.377 min
Delta R.T.: 0.018 min
Response: 865054
Conc: 89.84 ng/ml



#24 AR-1248-4


R.T.: 6.030 min
Delta R.T.: -0.027 min
Response: 14868910
Conc: 237.73 ng/ml

Instrument :
ECD_O
ClientSampleId :
CABLE-BRIDGE-FOUNDATION-D-5




#24 AR-1248-4

R.T.: 4.544 min
Delta R.T.: 0.030 min
Response: 2273264
Conc: 189.02 ng/ml




#25 AR-1248-5

R.T.: 6.097 min
Delta R.T.: 0.003 min
Response: 2035273
Conc: 33.64 ng/ml



#25 AR-1248-5


R.T.: 4.833 min
Delta R.T.: -0.040 min
Response: 3888527
Conc: 312.97 ng/ml



#26 AR-1254-1


R.T.: 6.030 min
Delta R.T.: 0.004 min
Response: 14868910
Conc: 94.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
CABLE-BRIDGE-FOUNDATION-D-5




#26 AR-1254-1

R.T.: 4.833 min
Delta R.T.: 0.000 min
Response: 3888527
Conc: 80.19 ng/ml




#27 AR-1254-2

R.T.: 6.242 min
Delta R.T.: 0.004 min
Response: 22067230
Conc: 89.98 ng/ml



#27 AR-1254-2


R.T.: 4.971 min
Delta R.T.: -0.002 min
Response: 2967164
Conc: 72.61 ng/ml



#28 AR-1254-3


R.T.: 6.602 min
Delta R.T.: 0.000 min
Response: 16094639
Conc: 63.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
CABLE-BRIDGE-FOUNDATION-D-5




#28 AR-1254-3

R.T.: 5.345 min
Delta R.T.: 0.002 min
Response: 8031903
Conc: 122.01 ng/ml




#29 AR-1254-4

R.T.: 6.903 min
Delta R.T.: 0.021 min
Response: 9094773
Conc: 48.22 ng/ml



#29 AR-1254-4


R.T.: 5.561 min
Delta R.T.: -0.002 min
Response: 1590298
Conc: 36.65 ng/ml



#30 AR-1254-5


R.T.: 7.290 min
Delta R.T.: -0.003 min
Response: 54807526
Conc: 306.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
CABLE-BRIDGE-FOUNDATION-D-5




#30 AR-1254-5

R.T.: 5.948 min
Delta R.T.: 0.000 min
Response: 13243045
Conc: 273.30 ng/ml




#31 AR-1260-1

R.T.: 6.752 min
Delta R.T.: -0.005 min
Response: 67871019
Conc: 969.65 ng/ml



#31 AR-1260-1


R.T.: 5.462 min
Delta R.T.: 0.000 min
Response: 14205710
Conc: 744.32 ng/ml



#32 AR-1260-2


R.T.: 7.009 min
Delta R.T.: -0.001 min
Response: 35080523
Conc: 405.63 ng/ml

Instrument :
ECD_O
ClientSampleId :
CABLE-BRIDGE-FOUNDATION-D-5




#32 AR-1260-2

R.T.: 5.644 min
Delta R.T.: -0.003 min
Response: 7811874
Conc: 334.34 ng/ml




#33 AR-1260-3

R.T.: 7.357 min
Delta R.T.: -0.002 min
Response: 23880781
Conc: 439.61 ng/ml



#33 AR-1260-3


R.T.: 5.778 min
Delta R.T.: -0.004 min
Response: 8238727
Conc: 381.26 ng/ml



#34 AR-1260-4


R.T.: 7.583 min
Delta R.T.: -0.002 min
Response: 27350363
Conc: 476.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
CABLE-BRIDGE-FOUNDATION-D-5




#34 AR-1260-4

R.T.: 6.223 min
Delta R.T.: -0.003 min
Response: 5396738
Conc: 386.75 ng/ml




#35 AR-1260-5

R.T.: 7.887 min
Delta R.T.: -0.003 min
Response: 55834356
Conc: 459.69 ng/ml



#35 AR-1260-5


R.T.: 6.464 min
Delta R.T.: -0.004 min
Response: 22504197
Conc: 663.49 ng/ml



#36 AR-1262-1


R.T.: 7.357 min
Delta R.T.: 0.000 min
Response: 23880781
Conc: 253.28 ng/ml

Instrument :
ECD_O
ClientSampleId :
CABLE-BRIDGE-FOUNDATION-D-5




#36 AR-1262-1

R.T.: 5.981 min
Delta R.T.: -0.002 min
Response: 6410212
Conc: 269.26 ng/ml




#37 AR-1262-2

R.T.: 7.887 min
Delta R.T.: 0.000 min
Response: 55834356
Conc: 352.68 ng/ml



#37 AR-1262-2


R.T.: 6.464 min
Delta R.T.: -0.002 min
Response: 22504197
Conc: 561.06 ng/ml



#38 AR-1262-3


R.T.: 8.181 min
Delta R.T.: 0.000 min
Response: 21562793
Conc: 649.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
CABLE-BRIDGE-FOUNDATION-D-5




#38 AR-1262-3

R.T.: 6.725 min
Delta R.T.: -0.002 min
Response: 4302663
Conc: 272.07 ng/ml




#39 AR-1262-4

R.T.: 8.235 min
Delta R.T.: -0.001 min
Response: 26567604
Conc: 332.47 ng/ml



#39 AR-1262-4


R.T.: 6.789 min
Delta R.T.: -0.002 min
Response: 11412768
Conc: 389.58 ng/ml



#40 AR-1262-5


R.T.: 8.810 min
Delta R.T.: 0.000 min
Response: 17609379
Conc: 328.73 ng/ml

Instrument :
ECD_O
ClientSampleId :
CABLE-BRIDGE-FOUNDATION-D-5




#40 AR-1262-5

R.T.: 7.272 min
Delta R.T.: -0.002 min
Response: 3982528
Conc: 290.54 ng/ml




#41 AR-1268-1

R.T.: 8.154 min
Delta R.T.: 0.000 min
Response: 16672956
Conc: 88.22 ng/ml



#41 AR-1268-1


R.T.: 6.725 min
Delta R.T.: -0.002 min
Response: 4302663
Conc: 94.34 ng/ml



#42 AR-1268-2


R.T.: 8.235 min
Delta R.T.: -0.004 min
Response: 26567604
Conc: 150.04 ng/ml

Instrument :
ECD_O
ClientSampleId :
CABLE-BRIDGE-FOUNDATION-D-5




#42 AR-1268-2

R.T.: 6.789 min
Delta R.T.: -0.004 min
Response: 11412768
Conc: 266.37 ng/ml




#43 AR-1268-3

R.T.: 8.425 min
Delta R.T.: 0.000 min
Response: 4400684
Conc: 28.01 ng/ml



#43 AR-1268-3


R.T.: 6.973 min
Delta R.T.: -0.001 min
Response: 1022112
Conc: 28.21 ng/ml



#44 AR-1268-4


R.T.: 8.810 min
Delta R.T.: 0.000 min
Response: 17609379
Conc: 295.14 ng/ml

Instrument :
ECD_O
ClientSampleId :
CABLE-BRIDGE-FOUNDATION-D-5




#44 AR-1268-4

R.T.: 7.272 min
Delta R.T.: -0.002 min
Response: 3982528
Conc: 262.20 ng/ml



#45 AR-1268-5

R.T.: 9.151 min
Delta R.T.: 0.001 min
Response: 22946334
Conc: 48.88 ng/ml



#45 AR-1268-5

R.T.: 7.526 min
Delta R.T.: -0.002 min
Response: 5261049
Conc: 48.93 ng/ml