

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0022420\
 Data File : P0066789.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Feb 2020 21:39
 Operator : DD\AJ
 Sample : L1660-01 10X
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 01-A-01-B-01-C

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 25 01:08:09 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022120.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 21 14:58:30 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.124	3.406	90816	114872	2.161	2.127
2) SA Decachlor...	9.596	8.289	162669	204013	3.964	3.671
Target Compounds						
26) L6 AR-1254-1	6.110	5.244	6743006	10155627	4254.103	3600.818
27) L6 AR-1254-2	6.327	5.390	17888337	14952161	6727.255	6224.180
28) L6 AR-1254-3	6.689	5.786	26618412	33033160	10066.735	8563.661
29) L6 AR-1254-4	6.973	6.014	28171874	23620155	12597.923	9861.124
30) L6 AR-1254-5	7.387	6.426	40691114	52400828	17622.929	14679.312

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0022420\
Data File : P0066789.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 Feb 2020 21:39
Operator : DD\AJ
Sample : L1660-01 10X
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
01-A-01-B-01-C

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 25 01:08:09 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022120.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 21 14:58:30 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

