

Data Path : P:\HPCHEM1\ECD_P\Data\PP022818\
 Data File : PP024054.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 Feb 2018 20:28
 Operator : UA/AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 01 03:15:29 2018
 Quant Method : P:\HPCHEM1\ECD_P\methods\PP020618.M
 Quant Title : 8080.M
 QLast Update : Tue Feb 06 02:07:55 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.573	11.363	367.0E6	439.5E6	716.137	724.704
Target Compounds							
1) T	Dalapon	3.013	2.580	476.3E6	479.7E6	596.759	579.514
2) T	3,5-DICHL...	9.982	9.481	516.8E6	546.4E6	649.048	655.072
3) T	4-Nitroph...	11.156	10.524	216.0E6	232.5E6	502.322	501.026
5) T	DICAMBA	11.891	11.660	1478.6E6	1587.0E6	687.616	698.791
6) T	MCPD	12.225	11.865	83559119	101.6E6	56.991	63.127
7) T	MCPA	12.480	12.257	126.4E6	146.3E6	57.377	62.807
8) T	DICHLORPROP	13.075	12.771	359.1E6	399.2E6	676.058	680.008
9) T	2,4-D	13.434	13.257	449.6E6	514.9E6	671.555	670.730
10) T	Pentachlo...	13.823	13.817	5694.0E6	6082.5E6	700.449	706.488
11) T	2,4,5-TP ...	14.687	14.413	2303.5E6	2502.1E6	695.413	699.379
12) T	2,4,5-T	15.086	15.045	2121.8E6	2048.4E6	679.108	664.944
13) T	2,4-DB	15.828	15.639	261.4E6	288.5E6	664.835	667.339
14) T	DINOSEB	17.294	16.040	1411.1E6	1463.9E6	614.790	595.089
15) T	Picloram	17.106	17.385	2959.3E6	2984.6E6	607.836	599.183
16) T	DCPA	17.662	17.261	2755.8E6	3014.0E6	702.805	692.713


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : P:\HPCHEM1\ECD_P\Data\PP022818\
Data File : PP024054.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 28 Feb 2018 20:28
Operator : UA/AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 01 03:15:29 2018
Quant Method : P:\HPCHEM1\ECD_P\methods\PP020618.M
Quant Title : 8080.M
QLast Update : Tue Feb 06 02:07:55 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 3.013 min
Delta R.T.: 0.000 min
Response: 476282039
Conc: 596.76 ng/ml

Instrument :
ECD_O
ClientSampleId :
HSTDCCC750




#1 Dalapon

R.T.: 2.580 min
Delta R.T.: 0.005 min
Response: 479667419
Conc: 579.51 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 9.982 min
Delta R.T.: -0.013 min
Response: 516837682
Conc: 649.05 ng/ml



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 9.481 min
Delta R.T.: -0.004 min
Response: 546383096
Conc: 655.07 ng/ml



#3 4-Nitrophenol


R.T.: 11.156 min
Delta R.T.: -0.013 min
Response: 215983707
Conc: 502.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
HSTDCCC750




#3 4-Nitrophenol

R.T.: 10.524 min
Delta R.T.: -0.004 min
Response: 232526086
Conc: 501.03 ng/ml




#4 2,4-DCAA

R.T.: 11.573 min
Delta R.T.: -0.016 min
Response: 366999404
Conc: 716.14 ng/ml



#4 2,4-DCAA


R.T.: 11.363 min
Delta R.T.: -0.006 min
Response: 439473438
Conc: 724.70 ng/ml



#5 DICAMBA


R.T.: 11.891 min
Delta R.T.: -0.016 min
Response: 1478632135
Conc: 687.62 ng/ml

Instrument :
ECD_O
ClientSampleId :
HSTDCCC750




#5 DICAMBA

R.T.: 11.660 min
Delta R.T.: -0.006 min
Response: 1586977404
Conc: 698.79 ng/ml




#6 MCP

R.T.: 12.225 min
Delta R.T.: -0.025 min
Response: 83559119
Conc: 56.99 ug/ml



#6 MCP


R.T.: 11.865 min
Delta R.T.: -0.013 min
Response: 101573307
Conc: 63.13 ug/ml



#7 MCPA


R.T.: 12.480 min
Delta R.T.: -0.027 min
Response: 126381317
Conc: 57.38 ug/ml

Instrument :
ECD_O
ClientSampleId :
HSTDCCC750




#7 MCPA

R.T.: 12.257 min
Delta R.T.: -0.017 min
Response: 146254678
Conc: 62.81 ug/ml




#8 DICHLORPROP

R.T.: 13.075 min
Delta R.T.: -0.018 min
Response: 359113576
Conc: 676.06 ng/ml



#8 DICHLORPROP


R.T.: 12.771 min
Delta R.T.: -0.007 min
Response: 399230348
Conc: 680.01 ng/ml



#9 2,4-D


R.T.: 13.434 min
Delta R.T.: -0.019 min
Response: 449620800
Conc: 671.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
HSTDCCC750




#9 2,4-D

R.T.: 13.257 min
Delta R.T.: -0.007 min
Response: 514936417
Conc: 670.73 ng/ml




#10 Pentachlorophenol

R.T.: 13.823 min
Delta R.T.: -0.015 min
Response: 5694003429
Conc: 700.45 ng/ml



#10 Pentachlorophenol


R.T.: 13.817 min
Delta R.T.: -0.006 min
Response: 6082507242
Conc: 706.49 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 14.687 min
Delta R.T.: -0.019 min
Response: 2303496005
Conc: 695.41 ng/ml

Instrument :
ECD_O
ClientSampleId :
HSTDCCC750




#11 2,4,5-TP (SILVEX)

R.T.: 14.413 min
Delta R.T.: -0.008 min
Response: 2502057901
Conc: 699.38 ng/ml




#12 2,4,5-T

R.T.: 15.086 min
Delta R.T.: -0.020 min
Response: 2121791540
Conc: 679.11 ng/ml



#12 2,4,5-T


R.T.: 15.045 min
Delta R.T.: -0.008 min
Response: 2048421174
Conc: 664.94 ng/ml



#13 2,4-DB


R.T.: 15.828 min
Delta R.T.: -0.024 min
Response: 261398403
Conc: 664.83 ng/ml

Instrument :
ECD_O
ClientSampleId :
HSTDCCC750




#13 2,4-DB

R.T.: 15.639 min
Delta R.T.: -0.009 min
Response: 288496218
Conc: 667.34 ng/ml




#14 DINOSEB

R.T.: 17.294 min
Delta R.T.: -0.025 min
Response: 1411053937
Conc: 614.79 ng/ml



#14 DINOSEB


R.T.: 16.040 min
Delta R.T.: -0.009 min
Response: 1463852256
Conc: 595.09 ng/ml



#15 Picloram


R.T.: 17.106 min
Delta R.T.: -0.026 min
Response: 2959347922
Conc: 607.84 ng/ml

Instrument :
ECD_O
ClientSampleId :
HSTDCCC750




#15 Picloram

R.T.: 17.385 min
Delta R.T.: -0.010 min
Response: 2984617410
Conc: 599.18 ng/ml



#16 DCPA

R.T.: 17.662 min
Delta R.T.: -0.028 min
Response: 2755775673
Conc: 702.80 ng/ml



#16 DCPA

R.T.: 17.261 min
Delta R.T.: -0.009 min
Response: 3013975974
Conc: 692.71 ng/ml