

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0030419\
 Data File : P0054211.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 04 Mar 2019 17:42
 Operator : SM/SJ
 Sample : AR1254ICC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 05 06:30:48 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0030419.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 05 06:30:24 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.034	3.118	210.5E6	11683593	50.000	50.000
2) SA Decachlor...	9.358	7.690	61258600	13069046	50.000	50.000
Target Compounds						
26) L6 AR-1254-1	5.988	4.805	39525669	8780614	500.000	500.000
27) L6 AR-1254-2	6.199	4.943	60560492	7327258	500.000	500.000
28) L6 AR-1254-3	6.562	5.312	60965710	12052530	500.000	500.000
29) L6 AR-1254-4	6.841	5.531	43933225	8003280	500.000	500.000
30) L6 AR-1254-5	7.251	5.917	39815521	8680478	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0030419\
Data File : P0054211.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 04 Mar 2019 17:42
Operator : SM/SJ
Sample : AR1254ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 05 06:30:48 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0030419.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 05 06:30:24 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.034 min
Delta R.T.: 0.000 min
Response: 210503595
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.118 min
Delta R.T.: 0.000 min
Response: 11683593
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.358 min
Delta R.T.: 0.000 min
Response: 61258600
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.690 min
Delta R.T.: 0.000 min
Response: 13069046
Conc: 50.00 ng/ml



#26 AR-1254-1


R.T.: 5.988 min
Delta R.T.: 0.000 min
Response: 39525669
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500




#26 AR-1254-1

R.T.: 4.805 min
Delta R.T.: 0.000 min
Response: 8780614
Conc: 500.00 ng/ml




#27 AR-1254-2

R.T.: 6.199 min
Delta R.T.: 0.000 min
Response: 60560492
Conc: 500.00 ng/ml



#27 AR-1254-2


R.T.: 4.943 min
Delta R.T.: 0.000 min
Response: 7327258
Conc: 500.00 ng/ml



#28 AR-1254-3


R.T.: 6.562 min
Delta R.T.: 0.000 min
Response: 60965710
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500




#28 AR-1254-3

R.T.: 5.312 min
Delta R.T.: 0.000 min
Response: 12052530
Conc: 500.00 ng/ml




#29 AR-1254-4

R.T.: 6.841 min
Delta R.T.: 0.000 min
Response: 43933225
Conc: 500.00 ng/ml



#29 AR-1254-4


R.T.: 5.531 min
Delta R.T.: 0.000 min
Response: 8003280
Conc: 500.00 ng/ml



#30 AR-1254-5

R.T.: 7.251 min
Delta R.T.: 0.000 min
Response: 39815521
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500



#30 AR-1254-5

R.T.: 5.917 min
Delta R.T.: 0.000 min
Response: 8680478
Conc: 500.00 ng/ml